

EXHIBIT 2

BRUCE ROSENZWEIG, M.D.'S GENERAL EXPERT REPORT

I. QUALIFICATIONS

I am currently an Assistant Professor of Obstetrics and Gynecology at Rush University Medical Center in Chicago, Illinois. I received my MD degree in 1984 from the University of Michigan in Ann Arbor, Michigan. Following graduation from medical school, I completed an Obstetrics and Gynecology Residency at Michael Reese Hospital in Chicago. In 1988, I attended a one year pelvic surgery fellowship at State University of New York in Syracuse, New York. Following that fellowship, I attended a two year Urogynecology and Urodynamics fellowship at UCLA Harbor General Hospital in Torrance, California. After graduating from the Urogynecology fellowship, I became a faculty member at the University of Illinois in Chicago. I started a Urogynecology program at the University of Illinois and also was the residency program director.

In 1998, I went into private practice, and subsequently established a private practice at Rush University Medical Center. I have also worked at John H. Stroger Hospital here in Chicago from May 2003 until November 2010 and Weiss Memorial Hospital as Associate Chair of Gynecology from February 2011 until July 2012. I have published numerous articles and given numerous lectures on the topics of pelvic organ prolapse, urinary incontinence and repair of pelvic organ prolapse.

Throughout my career, I have performed over a thousand pelvic floor surgical procedures, including abdominal sacrocolpopexy, uterosacral suspensions, sacrospinous ligament fixations, native tissue repairs, biological graft repairs and synthetic mesh repairs. I have also used numerous synthetic pelvic mesh products, including Ethicon's TVT, TVT-Obturator, and

Prolift. In addition, I have performed well over 200 surgeries dealing with complications related to synthetic mesh, including the removal of numerous TVT devices.

I was invited by Ethicon and attended both its Gynecare Prolift Training Seminar and TVT Obturator Seminar in Belgium. I have also attended a Bard Avaulta training seminar.

A copy of my CV and Fee Schedule is attached as Exhibit “A” and a copy of my testimony for the last four years is attached as Exhibit “B”. The documents I relied on for this report are contained in Exhibit “C” as well as those documents cited throughout this Report.

II. SUMMARY OF OPINIONS

In formulating my opinions and preparing this report, I reviewed scientific literature, corporate documents from Ethicon, regulatory documents in the public domain, sample products and depositions of Ethicon employees. A list of Ethicon corporate documents and depositions reviewed for this report is attached hereto as Exhibit “B”; all other materials reviewed are listed at the end of this report. All opinions I have are to a reasonable degree of medical and scientific certainty. I understand discovery is still ongoing in this case, and I reserve my right to amend my opinions if further information is provided in any form including, but not limited to, corporate documents, depositions and the expert reports of both Plaintiff and Defense experts.

In general, my expert opinions can be summarized as follows¹:

- A) Ethicon’s old construction TVT mesh (Prolene) is not suitable for its intended application as a permanent prosthetic implant for stress urinary incontinence because the pores are too small, it is heavy weight mesh, it degrades over time, causes chronic foreign body reactions, fibrotic bridging, mesh contracture/shrinkage, fraying, particle loss, sharp edges, roping and curling of the mesh, it deforms, and the pores collapse with tension;
- B) Ethicon knew that the old construction mesh (Prolene) was not appropriate for use in its TVT device as early as 1998, but has failed to modify/change the mesh to a larger pore, lighter weight mesh that would not deform, fray, lose particles, rope, curl,

¹ This is not intended to be an exhaustive recitation of my opinions in this case. The full scope of my opinions are described in further detail in this report.

- degrade, cause excessive foreign body reactions, and cause excessive shrinkage/contraction because of its economic interest in maintaining its competitive advantage in the MUS market and, therefore, Ethicon put profits before patient safety;
- C) Ethicon's warnings and disclosures of adverse events in its TVT Instructions for Use ("IFU") have been inadequate based on the adverse reactions and risks associated with the TVT that have been known to Ethicon from the time the TVT was first sold and marketed;
 - D) Ethicon did not disclose information to physicians in its IFUs regarding characteristics of the old construction TVT mesh (Prolene) that make it unsuitable for its intended application as a permanent prosthetic implant for stress urinary incontinence, including that it is small pore, heavy weight mesh, it degrades over time, causes chronic foreign body reactions, fibrotic bridging, mesh contracture/shrinkage, fraying, particle loss, roping and curling of the mesh., and that it deforms and the pores collapse with tension;
 - E) Ethicon failed to adequately describe, inform or explain to physicians how to properly "tension" the TVT and inform them that improper tension on the mesh decreased effective pore size and interfered with incorporation into tissue;
 - F) Ethicon did not inform physicians and patients that Manufacturer Safety Data Sheets (MSDSs) for polypropylene resin used to manufacture polypropylene meshes warned against use of the mesh in a permanently implanted medical device and that studies showed that it caused sarcomas in laboratory rats;
 - G) Ethicon did not properly inform physicians and patients that toxicity testing of the polypropylene mesh revealed that it was cytotoxic; ;
 - H) Ethicon's promotional materials sent to physicians related to TVT were inaccurate and failed to reveal material information about complications/risks and conflict of interests regarding data promoted in the materials;
 - I) Ethicon's Patient Brochures misstate and omit information regarding complications and success rates and overstate the benefits of the TVT while understating the risks.
 - J) Ethicon's Collection and Reporting of Adverse Events and Complications to Physicians and Patients Is Misleading, Inaccurate and Incomplete.
 - K) The Benefits of the TVT Are Outweighed By the Severe, Debilitating and Life Changing Complications Associated With the TVT

III. BACKGROUND AND TREATMENT OPTIONS FOR STRESS URINARY INCONTINENCE

A. Stress Urinary Incontinence²

Approximately one out of every three women over the age of 45 years old has some form of urinary incontinence. The majority of those women do not seek medical advice or treatment for a variety of reasons.

In a continent individual, increased abdominal pressure is evenly distributed over the bladder, bladder neck, and urethra. The urethral sphincter is thus able to withstand this pressure and maintain continence. In a person with pure stress urinary incontinence, either the urethra is hypermobile and/or the sphincter is intrinsically deficient. In urethral hypermobility, the urethrovesical junction (UVJ) is displaced extra-abdominally, and the increased intra-abdominal pressure is unevenly distributed such that the sphincter can no longer withstand the pressure and urine leaks. With intrinsic sphincter deficiency (ISD), the UVJ may not be hypermobile; however, the maximal urethral closing pressure, the Valsalva leak-point pressure, or both are too low to withstand the increase in intra-abdominal pressure and, thus, urine leaks past the sphincter.

Stress urinary incontinence (SUI) is the involuntary leakage of urine during moments of physical activity that increases abdominal pressure, such as coughing, sneezing, laughing, or exercise, in the absence of a bladder contraction. It has been estimated that 14% of women have SUI. SUI is a common type of urinary incontinence in women. Urodynamic proven SUI is found in approximately 50 % of women presenting for evaluation of urinary incontinence. Symptomatic women with SUI have social or hygienic consequence from their urine loss. SUI can happen when pelvic tissues and muscles, which support the bladder and urethra, become weak and allow the bladder “neck” (where the bladder and urethra intersect) to descend during

² For this section, see FDA.gov (2013) Stress Urinary Incontinence and Wikipedia.org (2013) Urinary Incontinence

bursts of physical activity (urethral hypermobility). This descent can prevent the urethra from working properly to control the flow of urine. SUI can also occur when the sphincter muscle that controls the urethra weakens (intrinsic sphincter deficiency). The weakened sphincter muscle is not able to stop the flow of urine under normal circumstances, and when there is an increase in abdominal pressure. Weakness may occur from pregnancy, childbirth, aging, or prior pelvic surgery. It has been estimate that a majority of women have a combination of urethral hypermobility and ISD. Other risk factors for SUI include chronic coughing or straining, constipation, obesity and smoking. Finally occult or latent SUI is defined as a positive stress test, loss of urine with increased intra-abdominal pressure and between 350-450cc volume in the bladder, after the repositioning of pelvic organ prolapse (usually accomplished with a ring pessary carefully positioned as to avoid compression of the urethra) in an otherwise clinically continent patient.

B. Nonsurgical Treatment of SUI.³

There are numerous non-surgical treatments available to woman with SUI. First, Pelvic Floor Exercises: A type of exercise to strengthen the pelvic floor by contracting and relaxing the levator muscles that surround the opening of the urethra, vagina, and rectum. These exercises, commonly referred to as Kegel exercises, improve the pelvic floor muscles' strength and function. Kegel exercises can improve over-active bladders by increasing urethral resistance with can trigger the bladder to relax.

Second, Pessary: A removable device that is inserted into the vagina against the vaginal wall and urethra to support the bladder neck. This helps reposition the urethra to reduce SUI. These can be made of rubber, latex or silicon. Inserted into the vagina, a pessary rests against the

³ For this section, see WebMD.com (2013) Mechanical Devices for Urinary Incontinence in Women and Netdoctor.com (2013) Stress Urinary Incontinence Pelvic Floor Exercises

back of the pubic bone and supports the bladder. Pessaries are available in various forms, including donut and cube shapes, and must be fitted by a healthcare provider. Some women who have stress incontinence use a pessary just during activities that are likely to cause urine leakage, such as jogging. Special incontinence pessaries have a 'knob', which fits under the urethra to elevate the midurethra to prevent urine loss.

Third, Transurethral Bulking Agents: Bulking agent injections are applied around the urethra that make the space around the urethra thicker, thus helping to control urine leakage. The effects are usually not permanent.

Fourth, Behavioral Modification: This includes avoiding activities that trigger episodes of leaking. Lifestyle modification can improve stress incontinence symptoms and include quitting smoking, weight loss, and allergy treatment during seasonal allergies.

Fifth, Urinary seals: These are adhesive foam pads, which women place over the urethral opening. The pad creates a seal and prevents the leakage of urine, providing incontinence treatment. The pad is removed before urination and replaced with a new one afterward. The pad can be worn during exercise or physical activity, but not during sexual intercourse.

Sixth, Urethral insert: A thin, flexible tube that is solid rather than hollow (like a catheter) is placed into the urethra to block the leakage of urine. These small plugs are inserted into the urethra by women to prevent leakage, and are removed prior to urination. These inserts can be uncomfortable and may increase the risk of urinary tract infection.

Seventh, Bladder neck support device: This device is a flexible ring with two ridges. Once inserted into the vagina, the ridges press against the vaginal walls and support the urethra. By lifting the bladder neck, it provides better bladder control in women suffering from stress incontinence. The device needs to be sized to fit, and must be removed and cleaned after

urination. Bladder neck support devices can be uncomfortable and may cause urinary tract infections.

C. Surgical Treatment of SUI

1. The Burch Colposuspension⁴

Retropubic approaches for the treatment of stress urinary incontinence include the Burch retropubic urethropexy (both open and laparoscopic) and the Marshall-Marchetti-Krantz (MMK) procedure. The goal of both of these procedures is to suspend and stabilize the urethra so that the urethrovesical junction (UVJ) and proximal urethra are replaced intra-abdominally and to recreate a firm backstop for intra-abdominal pressure. This anatomic placement allows normal pressure transmission during periods of increased intra-abdominal pressure restoring continence in a previously incontinent, hypermobile UVJ.

The Burch procedure was described in 1961. Initially, Burch described attaching the paravaginal fascia to the arcus tendineus. However, the point of attachment later changed to Cooper's ligaments because these were felt to provide more secure fixation points, and less chance of infection as seen with the prior MMK procedure.

Patients with type III stress urinary incontinence (a fixed, nonfunctioning proximal urethra) are not ideal candidates for a Burch procedure as no hypermobility exists to correct.

For the Burch procedure, a low Pfannestiel incision is made above the pubic bone in order to enter the space of Retzius (the anatomical space between the pubic bone and the bladder above the peritoneum in order to suspend the bladder and/or to perform a paravaginal repair. The procedure involves placing permanent stitches adjacent to the neck of the bladder and either proximal or distal to the bladder neck stitches on each side and suturing them to Cooper's

⁴ For this section, see Emedicine-Medscape.com (2013) Burch Colposuspension and Womensdoctor.com (2013) Burch Procedure and Paravaginal Repair.

ligament which is attached to the pubic bone. The paravaginal repair is very similar except that the stitches are attached to the arcus tendentious linea pelvis. The likelihood of success of the Burch and the paravaginal repair procedures is reported to be 80-90% in most cases. Success means total elimination of the incontinence and patient satisfaction score greater than 90%. Improved means significant reduction of urine loss and greater than 70% improvement of patient satisfaction scores. Additionally, these retropubic procedures can be accomplished by the laparoscopic route. With respect to the selection of synthetic absorbable suture versus non-absorbable suture, and braided versus monofilament, no prospective randomized blinded data exist to suggest superiority of one suture material over another. However, recognized risks are associated with bone anchors. Modifications in the technique can be used if co-existent central defect cystocele is present and obliteration of the cul-de-sac can be performed to prevent enterocele or posterior vaginal wall prolapse after Burch Colposuspension.

2. *Pubovaginal sling procedures*⁵

Pubovaginal slings have excellent overall success and durable cure. The procedure involves placing a band of autologous, allograft, xenograft or synthetic material directly under the bladder neck (i.e., proximal urethra) or mid-urethra, which acts as a physical support to prevent bladder neck and urethral descent during physical activity. This is brought up through the rectus fascia. The sling also may augment the resting urethral closure pressure with increases in intra-abdominal pressure.

Von Giordano is usually credited with performing the first pubovaginal sling operation in 1907, using a gracilis muscle graft around the urethra. In 1914, Frangenheim used rectus abdominis muscle and fascia for pubovaginal slings. In 1942, Aldridge, Millin, and Read

⁵ For this section, see Emedicine-Medscape.com (2013) Vaginal Sling Procedures and Womensdoctor.com (2013) Burch Procedure and Paravaginal Repair.

corrected urinary incontinence using fascial slings. In 1965, Zoedler and Boeminghous first introduced synthetic slings.

Historically, surgeons have used the fascia lata sling for recurrent SUI after a failed anti-incontinence operation. Furthermore, this operation is used extensively for the treatment of primary ISD. If the abdominal tissues are weak and attenuated or if the vaginal tissues are atrophied or in short supply, constructing a pubovaginal sling from the leg fascia lata can be performed. This procedure is more involved than the creation of the rectus fascial sling as it requires a second incision to harvest the fascia lata and healing in an area remote for the index procedure.

An alternative to a long rectus sling is construction of a short sling from a much smaller piece of abdominal fascia (rectus fascia suburethral sling). The surgical procedure is similar to that used for the rectus fascia pubovaginal sling, except that the harvested fascial tissue is much smaller and the operation time shorter. The advantage of this procedure is its simplicity. No extensive dissection in the suprapubic area is necessary, and the postoperative result is similar to that of the full-length fascial strip sling.

An alternative to a long fascia lata sling is the use of a postage stamp-sized patch of fascia lata from the outer thigh (fascia lata suburethral sling). The surgical procedure is similar to that for the fascia lata pubovaginal sling, except the harvested fascia is much smaller. This operation does not require extensive dissection in the thigh area, and the postoperative result is similar to that of the full-length fascia lata strip sling. Postoperative convalescence is shorter than that of the fascia lata pubovaginal sling procedure.

The vaginal wall suburethral sling helps restore urethral resistance by increasing urethral compression and improving mucosal coaptation of the bladder neck. This operation is attractive

because it is simple and easy to perform. Postoperative complications are minimal, and the recuperative period is short. Vaginal sling surgery is relatively contraindicated in elderly women with atrophic vaginitis. If recognized before surgery, the atrophied vaginal wall may be revitalized with the administration of vaginal estrogen cream or tablets for 3-6 months.

A clear contraindication to pubovaginal sling surgery is pure urge incontinence or mixed urinary incontinence (MUI) in which urge is the predominant component. An inherent risk of any sling procedure is de novo or worsening urge symptoms; thus, surgeons must identify and treat the presence of an urge component before surgery.

Conversely, poor detrusor function is a relative contraindication to pubovaginal sling surgery because the potential for urinary retention is increased. Women with absent or poor detrusor function in the presence of SUI are at a higher risk of experiencing prolonged postoperative urinary retention.

3. *Midurethral Synthetic Slings*

Based on the “Integral theory of female incontinence,” Prof. Ulmsten developed a midurethral procedure to treat stress urinary incontinence. The first reports of this procedure appeared in 1996 as an intravaginal slingplasty. The “tape” was placed through a small vaginal incision at the midurethra, brought through the urogenital diaphragm through the retropubic space and exited through small suprapubic incisions. The operation was theorized to correct incontinence by recreating the midurethral support of the pubourethral ligament and also by creating a midurethral hammock for support of the urethra during stress events. The procedure was described to have a success rate of 85-90% with an additional 5-10% significantly improved. The Gynecare TVT system was introduced in the US in November of 1998. Early studies

showed that the risk of bladder perforation during the procedure occurred 5-10% of cases and vascular injury with /without hematoma formation occurred in 2-5% of patients.

In an attempt to decrease the risk of bladder perforation and vascular injury, a ‘top-down’ approach to trocar placement was promoted as the SPARC procedure, introduced in the US in 2001 by American Medical Systems (AMS). The next modification of the midurethral sling came in 2001 when Delorme described his results for the use of the obturator membrane and inner thigh for passage of the sling material. The proposed advantage was avoidance of the retropubic space, thus avoiding bladder perforation and retropubic vascular injury. The trocars were passed from the inner thigh through the obturator membrane from an “outside – in direction”.

The next modification came from de Leval in 2003, with the “inside-out” trocar placement for the transobturator sling. The FDA’s Executive Summary 2011 found less mesh vaginal erosion in transobturator vs TVT (1.3%) vs (1.9%), less bladder perforation (0.3%) vs (5.5%), and less voiding dysfunction (4%) vs (7%), but a higher rate of groin pain (12%) vs (1.7%). Similar efficacy has been demonstrated for the transobturator vs retropubic approach.

The final modification came around 2006 with the release of the mini-slings, or single incision slings, which use support devices at the ends of shorter mesh lengths to accomplish fixation without the need for a secondary cutaneous exit point. The mini-slings could be placed in a retropubic or “U” fashion or a hammock or “H” fashion.

The FDA concluded in 2011 that there was higher peri-operative blood loss, higher mesh exposure and greater need for surgical re-intervention in the TVT-Secur (mini-sling) patients.

IV. EXPERT OPINIONS

- A. Ethicon's old construction TVT mesh (Prolene) is not suitable for its intended application as a permanent prosthetic implant for stress urinary incontinence because it is small pore, heavy weight mesh, it degrades over time, causes chronic foreign body reactions, fibrotic bridging, mesh contracture/shrinkage, fraying, particle loss, sharp edges, roping and curling of the mesh, and it deforms and the pores collapse with tension.**

Polypropylene mesh (Prolene), like that contained in the TVT, has many characteristics that make it unsuitable for use as a product intended for permanent implantation in the human vaginal floor. These characteristics include the following: (1) degradation of the mesh; (2) chronic foreign body reaction; (3) Infections and Bio-films; (4) fraying, roping, curling and deformation of the mesh; (5) loss of pore size with tension; (6) the combination of small pores and the heavy weight mesh causes fibrotic bridging leading to scar plate formation and mesh encapsulation; and (7) shrinkage/contraction of the encapsulated mesh.

As a result of these and other inadequacies with the mesh, and for the reasons set forth below, it is my opinion to a reasonable degree of medical certainty that the old construction Prolene polypropylene mesh in the TVT causes a multitude of injuries, including the possibility of multiple erosions that can occur throughout one's lifetime, chronic and debilitating pelvic pain, recurrence, worsening incontinence, chronic dyspareunia, wound infection, rejection of the mesh, sexual dysfunction, urinary and defecatory dysfunction, vaginal scarring, wound healing problems, injury to ureters, bladder and urethra, pelvic abscess formation, risk of infection, and/or the need for additional surgeries, among others. As a result, Ethicon's TVT mesh (Prolene) is not suitable for its intended application as a permanent prosthetic implant for stress urinary incontinence in women.

1. The Prolene Mesh in the TVT Degrades Over Time

As noted below, the mesh used in the TVT was originally designed in 1974 for use in the abdomen for treatment of hernias and it has not changed since then.⁶ Ethicon describes this mesh as the “old, old” mesh: “The first generation (old, old) mesh is utilized currently in the TVT product....”⁷ The current Material Specifications for TVT Mesh list it as: “Old Construction PROLENE* Mesh.”⁸ Dan Smith testified that even when the original hernia mesh was updated for use in the abdomen, Ethicon continued to use the “old, old” mesh for TVT and does to this day, as follows:

Q: So TVT kept the old when hernia changed to the new.

A: Also known as original, yes.

Q: The mesh that was used in the TVT-R is called sometimes by Ethicon in documents old construction or original mesh; correct?

A: Yes. Yes.⁹

In the late 90’s Ethicon determined that, in the hernia applications, it was safer to move to a lighter weight, larger pore mesh. Ethicon made a similar determination for meshes to be used in the pelvic floor.¹⁰ However, Ethicon never updated the “old, old” hernia mesh used in the TVT.¹¹ Notably, in my opinion this makes science and information regarding hernia meshes and other pelvis meshes of particular relevance when discussing the TVT mesh as Ethicon chose to move to large pore, light weight meshes in these areas, but not for TVT.

The placement of permanent polypropylene mesh in the human vagina creates problems because of the chemical composition and structure of the mesh and the physiological conditions

⁶ Smith Dep. (2/3/2014) 723:9-724:6.

⁷ Eth.Mesh.09275875.

⁸ Eth.Mesh.10633520 at 3522.

⁹ Smith Dep. (2/3/2014) 723:9-724:6.

¹⁰ See, e.g., Eth.Mesh.07455220 (discussing mesh shrinkage/contracture and stating: “Since this phenomenon occurs most frequently in small pore, heavy weight mesh, ETHICON has developed large pore, light weight meshes, i.e. GYNECARE GYNEMESH PS Nonabsorbable Prolene Soft Mesh....”).

¹¹ Smith Dep. (2/3/2014) 829:16-829:19.

of the vagina and the surrounding tissues. There have been numerous studies over the last 30 years which have shown polypropylene to be chemically reactive and not inert, with flaking and fissuring demonstrated by scanning electron microscopy, which leads to degradation and release of toxic compounds into pelvic tissues. This process enhances the inflammatory and fibrotic reactions within the tissues in the pelvic floor, causing a multitude of problems.¹² There have been recent studies suggesting that oxidation of the mesh occurs because of the polypropylene and the conditions in which it is placed.¹³ The oxidation causes the mesh to degrade, crack and break apart.¹⁴ In another recent study, 100 pelvic mesh implants were compared and over 20% showed degradation to mesh fibers.¹⁵

Because of the structural complexities of the vagina and the nature of the chemicals ordinarily found in the vagina and its surrounding tissues, there are several reasons why polypropylene presents unique problems when placed in the vagina. An Engineering Bulletin from Propex, entitled “*EB-405, The Durability of Polypropylene Geotextiles for Waste Containment Application*,” from 2011, states that, “[P]olypropylene is vulnerable to the following substances: highly oxidized substances such as (peroxide), certain chlorinated hydrocarbons (halogenated hydrocarbons), and certain aromatic hydrocarbons.”¹⁶ It is well known to physicians with expertise in the pelvic floor that vaginal and perivaginal tissues are ready sources for peroxide. The vaginal species lactobacillus produces hydrogen peroxide and

¹² Coda A, Hernia 2003;7:29; Jongbloed, WL, “Degradation of Polypropylene in the Human Eye: A SEM Study,” Doc Ophthalmol., 1986 64(1:143-152); Skrypunch, OW, “Giant Papillary Conjunctivitis from an Exposed Prolene Suture,” Can J Ophthalmology, 198621:(5: 189-192).

¹³ Costello C, Bachman S, Grant S, Cleveland D, Loy T, Ramshaw B, “Characterization of Heavyweight and Lightweight Polypropylene Prosthetic Mesh Explants from a Single Patient,” Surgical Innovation, 2007, 143: 168-176).

¹⁴ Id.

¹⁵ Clavé A, Yahi H, Hammou JC, Montanari S, Gounon P, Clavé H, “Polypropylene as a Reinforcement in Pelvic Surgery is Not Inert: Comparative Analysis of 100 Explants, : Int Urogynecol J 21:261-270 2007

¹⁶ Citing Schneider H., Long Term Performance of Polypropylene Geosynthetics, “Durability and Aging of Geosynthetics, Koerner, RM, Ed., (Elsevier 1989) 95-109.

lactic acid from glycogen that is produced in the squamous cells of the vagina. Estrogen is the catalyst for the production of glycogen from the vaginal cells. It is also well known that hydrogen peroxide produced by the lactobacillus species is important in controlling the vaginal micro-flora. In fact, the vagina is a ready source of hydrogen peroxide production. In a manuscript from M Strus, "*The In Vitro Effects of Hydrogen Peroxide on Vaginal Microbial Communities*," the authors show the amount of hydrogen peroxide produced by the lactobacillus species."¹⁷ "Hydrogen Peroxide reached concentrations of from 0.05 to 1.0 mM, which under intensive aeration increases even up to 1.8 mM."¹⁸ These results confirmed the previous results of M Strus in the publication, "*Hydrogen Peroxide Produced by Lactobacillus Species as a Regulatory Molecule for Vaginal Micro-flora*," Med Dosw Mikrobiol, 2004: 56(1:67-77).

It is also known that aromatic hydrocarbons can be found in the human body. In a paper from HB Moon entitled, "*Occurrence and Accumulation Patterns of Polycyclic Aromatic Hydrocarbons and Synthetic Musk Compounds in Adipose Tissues of Korean Females*," *Chemosphere* 2012 (86:485-490), these aromatic hydrocarbons were noted to be present in, "[t]otal concentrations of PAHs and SMCs in adipose tissues rang[ing] from 15 to 361 (mean:119) ngg(-1) lipid weight and from 38 to 253 (mean:106) nng(-1) lipid weight respectively The results of this study provide baseline information on exposure of PAHs and SMCs to the general population in Koreans."

It has also been determined that halogenated hydrocarbons can be found not only in adipose tissue but also the blood stream. A paper entitled, "*Determination of Volatile Purgeable Halogenated Hydrocarbon in Human Adipose Tissue and Blood Stream*," from the *Bulletin of Environmental Contamination and Toxicology*, Volume 23, Issue 1, pp 244 – 249 published in

¹⁷ Strus, M, et al., "The In Vitro Effect of Hydrngen Peroxide in Vaginal Microbial Communities," FEMS Immunol Med Microbiol, 2006 Oct; 48(1:56-63).

¹⁸ Id.

1979, found halogenated hydrocarbons, pesticide by-products, both in human adipose tissues and the blood stream. In a subsequent paper from 1985 in *Environmental Health Perspectives*, Volume 60, pp. 127-131, Henry Anderson, in his paper entitled, "*Utilization of Adipose Tissue Biopsy and Characterizing Human Halogenated Hydrocarbon Exposure*," also found these pesticide by-products in human adipose tissue. Accordingly, the body location where the polypropylene mesh is being placed can expose it to known chemical degradation agents.

However, chemical degradation is not the only way that polypropylene degrades in vivo. In a paper from N Das in the *Journal of Biotechnology Research International*, Volume 2011, Article ID 941810, entitled, "*Review Article: Microbial Degradation of Petroleum Hydrocarbons Contaminant: An Overview*," found that various bacteria such as *Pseudomonas* species, *Bacillus* species, *Mycobacterium* and *Corynebacterium* species, which are present in a woman's vagina, can degrade petroleum hydrocarbons. Also fungi such as the *Candida* species, also present, can degrade petroleum-based hydrocarbons.¹⁹

Microbial agents that can be found inside the normal and abnormal flora of the human vagina such as *Candida* and, with certain pelvic infections such as *Bacillus* and *Pseudomonas*, can be a source of biological degradation of polypropylene products. A paper entitled, "*Health, Safety and Environment Fact Sheet: Hazardous Substances - Plastics*," from CAW/TCA (www.caw.ca), August 2011:343, found that polypropylene degradation products and residues can form carbon monoxide, acrolein, aldehydes and acids, qualifying these health hazards as toxic and irritants. In a paper from D Lithner in 2011 at 4, entitled, "*Environmental and Health Hazards of Chemicals in Plastic Polymers and Products*", University of Gothenburg, it is stated that, "[n]on-biodegradable polymers can be degraded by heat, oxidation, light, ionic radiation,

¹⁹ Das, N , et al., "Review Article: Microbial Degradation of Petroleum Hydrocarbon Contaminants: an Overview," *J Biotech Res Intl*, 2011, Article ID 941810, 1-13.

hydrolysis and mechanical shear, and by pollutants such as carbon monoxide, sulphur dioxide, nitrogen oxide and ozone. This causes the polymer to get brittle, to fragment into small pieces and to release degradation products.” (Citations omitted.) She continues, “[o]ther substances (besides monomers) are often needed for polymerization to occur, for instance initiators, catalysts, and, depending on manufacturing process, solvents may also be used. The resulting plastic polymer can be blended with different additives, for instance plasticizers, flame retardants, heat stabilizers, antioxidants, light stabilizers, lubricants, acid scavengers, antimicrobial agents, anti-static agents, pigments, blowing agents and fillers, and is finally processed into a plastic product. There are many different plastic polymers and several thousand different additives, which result in an extremely large variation in chemical composition of plastic products.”²⁰ “Since plastic products are composed of many different chemicals, and the main part of these [are] broken down into something completely different; this complicates the prediction.” *Id.* at 8. “The type and quantity of degradation products formed may also be influenced by degradation mechanisms, presence of polymerization impurities, and surrounding factors, e.g. temperature and oxygen.” *Id.* at 9. “Few studies combining leaching tests with toxicity tests have been performed on plastic products.” *Id.* at 12.

The available peer-reviewed literature regarding degradation/oxidation of polypropylene in the human body dates back to the 1960’s and has been reported in numerous such publications.²¹ Two of the more important and salient articles regarding reported degradation in explanted surgical meshes (hernia and pelvic floor) are the Costello and Clave articles.

²⁰ *Id.* at 6 (citations omitted).

²¹ Liebert T, Chartoff R, Costgrove S, “Subcutaneous Implants of Polypropylene Filaments,” J Biomed Mater Res. 1976 (10:939-951); Williams D, “Review of Biodegradation of Surgical Polymers,” J Materials Sci. 1982 (17:1233-1246); Oswald, HJ, Turi, E, “The Deterioration of Polypropylene By Oxidative Degradation,” Polymer Eng Sci. 1965 (5:152-158).

In his paper, “*Characterization of Heavyweight and Lightweight Polypropylene Prosthetic Implants from a Single Patient*,” Prof. C Costello reported that hernia mesh made of polypropylene oxidized and degraded as a result of the metabolites produced by phagocytic cells during the body’s inflammatory reaction to the mesh. High-magnification photographs showed cracking and peeling of the polypropylene fibers. Ethicon was aware of this article as referenced in internal emails.²²

Another article by A Clave, “*Polypropylene as a Reinforcement in Pelvic Surgery is Not Inert: Comparative Analysis of 100 Explants*,” also displayed high magnification photos of polypropylene fibers from explanted meshes and, in this case, the meshes were explanted from women’s pelvic floor tissue.²³ The heavyweight meshes showed even greater cracking than the lower density meshes, but according to Prof/Dr. Clave, ALL 84 of the polypropylene explants examined showed degradation. Oxidation of the implanted mesh due to free radical attack through the synthesis of peroxides, superoxides and hypochlorous acid during the chronic inflammatory phase was listed as just one potential cause for the oxidative degradation within the “septic environment” in which the pelvic meshes are placed.

Given the information available to Ethicon in the scientific and medical literature concerning the potential for degradation of polypropylene, it is my opinion to a reasonable degree of medical certainty that Ethicon should have conducted clinically relevant testing to determine if naturally occurring conditions in the vagina could cause polypropylene to degrade and if so, what the quantity and quality of the products of degradation would be, whether they would be released into surrounding tissues and/or migrate in the woman’s body, what the clinical

²² Eth.Mesh.005588123

²³ Clave, A., Polypropylene as a Reinforcement in Pelvic Surgery is Not Inert: Comparative Analysis of 100 Explant, I Urogynecol J, 2010 21:261-270.

implications for the woman would be and whether some women's body's would react differently to the mesh and the degradative process and its by-products.

Daniel Burkley, a Principal Scientist at Ethicon, testified that the science supports the conclusion that mesh can shrink, contract and degrade. Specifically, Mr. Burkley agreed that the risk of degradation increases when you have a severe inflammatory response with mesh implanted in a contaminated field.²⁴ Mr. Burkley also testified that polypropylene mesh in human beings is subject to some slight degree of surface degradation.²⁵ He agreed that degradation might be better understood if Ethicon studied or tested a product that is permanently implanted in women.²⁶ In fact, according to Mr. Burkley, Ethicon only conducted one study related to degradation and Prolene material. This study consisted of a Prolene suture implanted into dogs.²⁷ Mr. Burkley testified that the study and photos from the dog actually showed that the Prolene material used in TVT degraded and was still degrading after 7 years.²⁸

It is now clear from Ethicon's internal documents that Mr. Burkely was incorrect when he said that Ethicon only performed one study related to degradation of Prolene. Contrary to Mr. Burkley's claim, he and other Ethicon scientists were involved in a Prolene human explant study that was conducted in 1987 – 10 years before TVT was marketed – which found that Prolene degrades while in the body. According to Ethicon's documents, Ethicon's scientists received 58 Prolene human explants from Professor Robert Guidoin²⁹ which were analyzed by Ethicon's scientists using scanning electron microscopy (SEM). The SEM study revealed that 34 of the 58 Prolene explants (58%) were cracked. Further studies, including FTIR and melt point analysis, were conducted by Ethicon's scientists to determine the cause of the cracking observed in Prof.

²⁴ Burkley 5/22/13 184:17-24.

²⁵ Burkley 5/22/13 206:2-11.

²⁶ Burkley 5/22/13 206:12-25.

²⁷ Eth.Mesh.05453719 (Seven year data for ten year Prolene study: ERF 85-219)

²⁸ Burkley 5/23/13 315:8-13.

²⁹ DEPO.ETH.MESH.00001766

Guidoin's explants. In a report authored Mr. Burkley on September 30, 1987, he concluded that the Prolene explants had insufficient antioxidants to protect them from oxidation which led to *in vivo* degradation of the Prolene devices.³⁰ Importantly, Ethicon has not made any changes to Prolene since it was introduced to the market, except that, in 1991, they reduced the amount of Santanox (another antioxidant), which could potentially make Prolene more, not less, susceptible to oxidized degradation.³¹ Thus, Ethicon's internal studies which predate the marketing of TVT by at least 10 years clearly demonstrate that Ethicon's scientists had concluded that Prolene can degrade while implanted in the human body.

Ethicon subsequently hired an outside consulting firm to resolve the cause of the erosion of its surgical meshes for the pelvic floor. In a June 22, 2011 report, PA Consulting Group informed Ethicon that, "[p]olypropylene can suffer from degradation following implant . . . a process which initiates after a few days post implantation in animal studies."³² The consulting report discusses numerous images of polypropylene mesh that show "physical degradation" of the mesh.³³ In addition, in a 2009 presentation, Ethicon Medical Director, Dr. Piet Hinoul, stated that meshes are not biologically inert.³⁴

I have personally seen mesh that is broken, cracked, deformed and looks degraded during explant surgeries. Interestingly, despite years of the scientific literature, its own internal studies and reports from consultants it hired that state degradation of the mesh occurs, Ethicon's Instructions for Use (IFU) continues to claim to this day that the mesh in the TVT, "is not

³⁰ ETH.MESH.12831391 at ETH.MESH.12831392

³¹ ETH.MESH.02268619

³² Eth.Mesh.02589032 and Eth.Mesh.07192929 (May 18, 2011 PA Consulting Report: Investigating Mesh Erosion in Pelvic Floor Repair and PowerPoint presentation).

³³ *Id.*

³⁴ Eth.Mesh.01264260 (Presentation, "Prolift+M," P Hinoul, MD, Ethicon Pelvic Floor Expert's Meeting – Nederland, Utrecht, May 7, 2009).

absorbed, nor is it subject to degradation or weakening by the action of enzymes.”³⁵ This is not simply inaccurate, but is false and misleading for all of the reasons stated above, including, most importantly, that Ethicon’s own internal documents and testimony from its employees confirm that the TVT degrades. Furthermore, Thomas Barbolt, Ethicon’s own scientist who was designated by Ethicon as the Ethicon scientist knowledgeable about Ethicon’s degradation studies and the statements Ethicon makes in its IFU about degradation, testified that Ethicon knew as early as the early 1990s that Prolene degrades, well before Ethicon decided to make the statement in the IFU that the Prolene is **not** subject to degradation.³⁶ Ethicon’s actions in this regard are clearly contrary to a medical device manufacturer’s good faith duty to honestly and truthfully inform physicians about its products.

It is my opinion to a reasonable degree of medical certainty that the mesh in the TVT degrades. The effect of chemical and biological degradation of the TVT Prolene mesh in a woman’s tissues can lead to a greater foreign body reaction, enhanced inflammatory response and excessive scarring, which can lead to severe complications in patients, including the possibility of multiple erosions that can occur throughout one’s lifetime, chronic and debilitating pelvic pain, recurrence, worsening incontinence, chronic dyspareunia, wound infection, rejection of the mesh, sexual dysfunction, urinary and defecatory dysfunction, vaginal scarring, wound healing problems, injury to ureters, bladder and urethra, pelvic abscess formation, risk of infection, and/or the need for additional surgeries, among others. As a result, the polypropylene in Ethicon’s TVT mesh (Prolene) is not suitable for its intended application as a permanent prosthetic implant for stress urinary incontinence in women.

³⁵ Eth.Mesh.05225354; Eth.Mesh.02340306; Eth.Mesh.02340471; Eth.Mesh.05222673; Eth.Mesh.02340504; Eth.Mesh.03427878.

³⁶ Deposition of Thomas Barbolt, M.D., January 8 2014, pg 409; 516-17

Given the information available in the scientific and medical literature concerning the potential for degradation of polypropylene, it is my opinion to a reasonable degree of medical certainty that Ethicon should have conducted clinically relevant testing to determine if naturally occurring conditions in the vagina could cause polypropylene to degrade and if so, what the quantity and quality of the products of degradation would be, whether they would be released into surrounding tissues and/or migrate in the woman's body, what the clinical implications for the woman would be, and whether some women's body's would react differently to the mesh and the degradative process and its by-products.

2. Chronic Foreign Body Reaction

The human body has a natural and fairly predictable "host defense response" to any foreign object placed inside of it. Whether a splinter or a surgical mesh, the human body will send white blood cells to attack the invader and, if the products of inflammation cannot ward off or destroy the invader, including the invader is anything from bacteria to prosthetic implants, the initial acute inflammatory phases is followed by a chronic inflammatory phase. Therefore, with the placement of something like a permanent surgical mesh in human tissues, there will be a chronic or permanent foreign body reaction to the implant, as well as a chronic inflammatory response by the body.³⁷ In fact, Ethicon Medical Directors, Dr. Piet Hinoul and Dr. Charlotte Owens, have both testified that the chronic foreign body reaction created by the body's response

³⁷ Klinge U, Klosterhalfen B, Muller M, Ottinger A, Schumpelick V, "Shrinking of Polypropylene Mesh In Vivo: An Experimental Study in Dogs," Eur J Surg 1998 (164: 965-969); Klinge U, Klosterhalfen B, Muller M, Schumpelick V, "Foreign Body reaction to Meshes Used for the Repair of Abdominal Wall Hernias," Eur J Surg, 1998 (164:951-960); Klosterhalfen, B, Junge, K, Klinge, U, "The lightweight and large porous mesh concept for hernia repair," Expert Rev. Med. Devices, 2005 2(1); Binnebosel M, von Trotha K, Jansen P, Conze J, Neumann U, Junge K, "Biocompatibility of prosthetic meshes in abdominal surgery" Semin Immunopathol, 2011 (33:235-243); Eth.Mesh.03658577 (Biocompatibility of Ultrapro).

to mesh can cause a severe inflammatory reaction, which can cause chronic pain, nerve entrapment, erosions, dyspareunia and the need for additional surgeries.³⁸

Ethicon was informed by its consultants and experts in the field that there will be chronic tissue reaction to its polypropylene meshes. During a 2006 meeting at one of Ethicon's facilities, Bernd Klosterhalfen, a pathology consultant expert for Ethicon, informed Ethicon that there can be a continuing reaction between tissues in the body and mesh for up to 20 years.³⁹ In addition, during a February 2007 meeting, Ethicon stated that there can be, "[E]xcessive FBR [foreign body reaction]> massive scar plate > more shrinkage."⁴⁰

Internally, Ethicon's scientists agreed. Dr. Holste testified that chronic foreign body reactions occur in Ethicon's small pore, heavyweight meshes like the Prolene mesh found in the TVT. In fact, Dr. Holste testified that Ethicon developed lighter weight, large pore meshes in order to minimize the complications seen with heavyweight meshes like the Prolene used in TVT.⁴¹ Ethicon employee, Christophe Vailhe, testified that there can be an excessive inflammatory reaction or foreign body reaction that would lead to mesh erosion and contraction.⁴² Despite its knowledge about the problems associated with the chronic foreign body reaction after implantation of the TVT, Ethicon continues to use a heavy weight, small pore mesh in its TVT product.

Contrary to this scientific evidence, Ethicon informed doctors in its IFU that its TVT mesh was "non-reactive with a minimal and transient foreign body reaction."⁴³ This was despite all of the internal documents and testimony discussed above from Ethicon's Medical Affairs and

³⁸ Hinoul 4/5/12 99:09-25; 4/6/12 518:14-520:20; 6/26/13 175:1-176:17,;184:18-22; 328:10-24; Owens 9/12/2012 98:11-99:07.

³⁹ Eth.Mesh.00870466 (June 6, 2006 Ethicon Expert Meeting Meshes for Pelvic Floor Repair, Norderstedt).

⁴⁰ Eth.Mesh.01218361 (Ethicon Presentation: "State of Knowledge in 'mesh shrinkage'-What do we know").

⁴¹ Holste 7/29/13 51:3-53:6.

⁴² Vailhe 6/21/13 383:8-19.

⁴³ Eth.Mesh.05225354; Eth.Mesh.02340306; Eth.Mesh.02340471; Eth.Mesh.05222673; Eth.Mesh.02340504; Eth.Mesh.03427878.

Research and Development employees that chronic foreign body reaction occurs in small pore, heavyweight meshes like the Prolene mesh in the TVT. Moreover, as one of Ethicon's lead engineers stated: "the foreign body reaction is not transitory – it doesn't ever go away, but decreases over time to a minimal level."⁴⁴ That is, it is chronic. I have reviewed numerous pathology reports from my own patients and other physician's patients and pathology reports reviewed in litigations describing foreign body reactions. Hence, the mesh potentiates a chronic, long-term inflammation. This is contrary to the express language of the TVT IFU and, to this date, has yet to be corrected in the IFU.

For the reasons set forth above, it is my opinion to a reasonable degree of medical certainty that the Prolene polypropylene mesh in the TVT creates a chronic foreign body reaction which can lead to severe complications in patients, including the possibility of multiple erosions that can occur throughout one's lifetime, chronic and debilitating pelvic pain, recurrence, worsening incontinence, dyspareunia that can be chronic, wound infection, rejection of the mesh, sexual dysfunction, urinary and defecatory dysfunction, vaginal scarring, wound healing problems, injury to ureters, bladder and urethra, pelvic abscess formation, risk of infection, and/or the need for additional surgeries, among others. As a result, the polypropylene in Ethicon's TVT mesh (Prolene) is not suitable for its intended application as a permanent prosthetic implant for stress urinary incontinence in women.

3. *Infections/Bio-films*

The placement of midurethral slings, including the TVT, violates one of the most basic tenets of surgical teachings in that it is the placement of a permanent implant into the human through a "clean contaminated" surgical field, i.e. the vagina, which is not sterile and can never

⁴⁴ Eth.Mesh.00211259

be completely sterilized. Therefore, implantation through the vagina is contraindicated for every procedure and implantation.

The weave of the mesh in the TVT produces very small interstices which allow bacteria to enter and to hide from the host defenses designed to eliminate them. The bacteria can secrete an encasing polysaccharide slime (biofilm), which further serves to shield the bacteria from destruction by white blood cells and macrophages.⁴⁵ The effect and consequences of biofilm to increase the foreign body reaction, resulting in chronic infections, chronic inflammation, erosions, and mesh and scar contracture, was well known to Ethicon, as evidenced by the testimony of Ethicon's Head of Pre-Clinical, Dr. Jorge Holste.⁴⁶ Importantly, the biofilm actually serves as a protection for the bacteria surrounding the mesh fibers against the body's host defense response (white blood cells), which are intended to destroy foreign invaders like bacteria. Thus, the weave induces the creation of a shield against the body's defenses to the bacteria entrained in the woven mesh, inhibiting the body's ability to fight off the infective agents within the mesh.

The large surface area promotes wicking of fluids and bacteria and is a "bacterial super highway" which provides a safe haven for bacteria which attach themselves to the mesh during the insertion process.⁴⁷ Daniel Burkley testified that reducing surface area could reduce the amount of chronic inflammation.⁴⁸ Additionally, the size of the mesh placed equates to a large

⁴⁵ Osterberg B, *et al.*, "Effect of Suture Materials on Bacterial Survival in Infected Wounds: An Experimental Study," *Acta. Chir. Scand* 1979 (145:7 431-434); Merritt K, "Factors Influencing Bacterial Adherence to Biomaterials," *J Biomat Appl* 1991 (5:185-203); An, Y, "Concise Review of Mechanisms of Bacterial Adhesion to Biomaterial Surfaces," *J Biomed Mater Res (Appl Biomat)*, 1998 (43:338-348); The TVM Group: J. Berrocal, *et al.* Conceptual advances in the surgical management of genital prolapsed, *J Gynecol Obstet Biol Reprod* 2004: 33:577-587.

⁴⁶ Holste 7/30/13 295:24-298:14, 411:15-414:24.

⁴⁷ Klinge, U, *et al.*, "Do Multifilament Alloplastic Meshes Increase the Infection Rate? Analysis of the Polymeric Surface, the Bacteria Adherence, and the In Vivo Consequences in a Rat Model," *J Biomed Mater Res*, 2002 (63:765-771); Vollebregt, A, *et al.*, "Bacterial Colonisation of Collagen-Coated Polypropylene Vaginal Mesh: Are Additional Intraoperative Sterility Procedures Useful?" *Int Urogyn J*, 2009 (20:1345-51).

⁴⁸ Burkley 5/22/2013, 371.

surface area with many places for bacteria to hide while being protected from host defenses leading to numerous complications.⁴⁹

There have been numerous peer-reviewed journal articles regarding secondary-mesh related infections as well as the dangers of implanting surgical mesh in a clean/contaminated field. Of note, in May of last year, at the AUA meeting in San Diego, Dr. Shah and his colleagues reported on the “*Bacteriological Analysis of Explanted Transvaginal Meshes*,” which included explanted samples of both SUI slings and prolapse meshes. Of the 50 explants examined, 52% of those explanted due to patient complaints’ of painful mesh were infused with pathogenic organisms, 20% of those explanted due to vaginal erosions had pathogenic organism, and 83% of those explanted due to urinary tract erosions were contaminated with pathogenic organisms.⁵⁰

When polypropylene particles separate from the surface of the mesh fiber due to degradation, see *infra*, the surface area of the mesh is greatly increased thus providing even greater areas for bacterial adherence to the mesh, more elution of toxic compounds from the polypropylene, and also more of the free toxic polypropylene itself, all of which increases the inflammatory reaction and intensity of the fibrosis.⁵¹ This cracking of the mesh surface also provides safe harbors for infectious bacteria to proliferate.

In his periodic histopathological analyses for Ethicon of its pelvic floor explants, Dr. Klosterhalfen reported to Ethicon that, in virtually 100% of those instances in which mesh had been explanted due to erosions, he found a secondary, mesh-related infection at the tissue/mesh

⁴⁹ Klinge, *supra*, at n 26; Vollebregt, *supra*, n. 26.

⁵⁰ Shah, K., et al., Bacteriological Analysis of Explanted Transvaginal Meshes (Abstract 1144)

⁵¹ Jongebloed, *supra*, n. 1; Sternschuss, G, Ostergard, DR, Patel H, “Post-Implantation Alterations of Polypropylene in the Human,” *J Urology*, 2012 (188:27-32); Clave, *supra*, at 6.

interface.⁵² In his analyses, Dr. Klosterhalfen also discusses neuromas and neuronal proliferation being found in a periphery of pelvic floor products which induces chronic pelvic pain. Mesh exposure and erosion cause the fibers to be further exposed to bacteria that will adhere to and colonize on the mesh surface.⁵³

Ethicon employees have testified that they were aware of these biofilms forming on the surface of the mesh.⁵⁴ However, Ethicon never performed any long-term clinical studies to determine whether the warnings given them through the peer-reviewed literature and by their own experts and consultants were accurate, namely that mesh-related infections are real; that they cause patient injury in the form of mesh erosions and recurrent, late infections; and that the transvaginal implantation through and into the non-sterile, contaminated vagina is below the standard of care for any surgical technique, especially one used to treat non-life threatening conditions, such as stress urinary incontinence.

Therefore, it is my opinion to a reasonable degree of medical certainty that the TVT mesh is susceptible to biofilm formation due to the weave of the mesh allowing the infiltration, harboring, and protection of bacterial contaminants; the degraded mesh surface harboring bacteria, the passage through and into a clean/contaminated field, and after exposure/erosion of the mesh into the vagina or other organs, further contamination of the mesh with a multitude of vaginal flora occurs that further increases the risk of harmful and recurrent infections in women.

Accordingly, the TVT transvaginal technique, as well as the TVT mesh itself, are not safe for their intended purpose of implantation into a woman's pelvic tissues and can lead to severe complications in patients, including the possibility of multiple erosions that can occur throughout

⁵² Eth.Mesh. 00006636

⁵³ Eth.Mesh. 00006636

⁵⁴ Holste, 7/30/13, 283:19-284:5

one's lifetime, chronic and debilitating pelvic pain, recurrence, worsening incontinence, dyspareunia that can be chronic, wound infection, rejection of the mesh, sexual dysfunction, urinary and defecatory dysfunction, vaginal scarring, wound healing problems, injury to ureters, bladder and urethra, pelvic abscess formation, risk of infection, and/or the need for additional surgeries, among others. As a result, Ethicon failed to act as a reasonable and prudent medical device manufacturer by manufacturing and selling its Prolene mesh in a permanent prosthetic implant like the TVT.

4. *Fibrotic Bridging*

Fibrotic bridging occurs when the fibers surrounding the pores of the mesh are too close together to allow the tissue in the pore enough room to recover from the trauma of tissue damage due to implanting a surgical prosthetic device. Pores that are large enough for good, newly-vascularized tissue tend to be filled with fatty tissue versus small pores that become filled with scarred or fibrotic tissue. In those instances, the scar forms across the pores or "bridges" from one side of the pore to the other. This can occur either due to the granulomas around the mesh fibers joining together or due to densely-formed fibroblast plugs between these granulomas. Either way, such bridging can lead to the creation of a rigid, scar plate that can encapsulate the mesh with scar tissue. Simply put, small mesh pores that cause fibrotic bridging turn the mesh into a solid sheet of scar tissue and there is no space or room for tissue to grow into the mesh, which is the intended purpose of the mesh. The fibrotic bridging and scar plate prevents tissue in-growth and causes complications, including, among other things, pain with the rigid mesh, shrinkage, contraction of the mesh, erosions due to mechanical irritation in the tissue of a rigid, scar-plated mesh, nerve entrapment, chronic pain and dyspareunia.

It is clear from Ethicon internal documents that it was well aware of fibrotic bridging.⁵⁵ Ethicon employees have testified that the heavy weight, small pore type of mesh in the TVT can lead to an increased risk of foreign body reaction, contraction of the mesh, nerve entrapment, erosions and chronic pelvic pain.⁵⁶ The problems with small pore, heavy weight mesh are best illustrated by a DVD produced by Ethicon which features an Ethicon consultant, Dr. Todd Heniford, talking about a heavy weight, small pore mesh called Marlex used for hernia repairs⁵⁷ The Prolene mesh used in TVT is of heavyweight, small pore construction and, in fact, is even heavier than Marlex. Ethicon Scientists have acknowledged the video and that the Marlex mesh is similar to the Prolene in TVT in that it is heavy weight small pore mesh.⁵⁸ In the video, Dr. Heniford talks about the dangers of heavy weight, small pore mesh.⁵⁹ In fact, Dr. Heniford states, “there is no excuse for using heavy weight, small pore meshes in the human body”.⁶⁰ I have explanted numerous TVT meshes and have witnessed meshes with extensive scar plating and mesh encapsulation similar to the hardened/stiffened mesh viewed in the Heniford video.

⁵⁵ Eth.Mesh.04037600 Innovations in mesh development; Eth.Mesh.05920616 7/20/07 ;Emails from Chomiak, Martin to Batke, Boris; Jamieson, Gillian; Koehler, Petra; and Hellhammer, Dr. Brigitte, SUBJECT: Defining light weight mesh; TH.MESH.05585033 Boris Batke Presentation – Project Edelweis - Ultrapro; Eth.Mesh.05446127 3/13/2006 Emails from Holste, Dr. Joerg to Engel, Dr. Dieter; Manley, Quentin; Storch, Mark L. SUBJECT: AW: Mesh and Tissue Contraction in Animal – “Shrinking Meshes? – Scientific Statement by Ethicon GmbH, R&D Europe; Biocompatibility of Meshes by Dr. J. Holste; Eth.Mesh.05475773 2/09/2007 Boris Batke, Ethicon R&D, Presentation: The (clinical) argument of lightweight mesh in abdominal surgery; Eth.Mesh.04015102 3/01/12 Email from Batke, Boris to Mayes, Casey SUBJECT: AW: AGES Pelvic Floor Conference-Gala Dinner Invitation; Eth.Mesh.04037600 3/15/12 Boris, Batke PowerPoint Presentation, Innovations in Mesh Development, Melbourne AGES 2012.

⁵⁶ Batke 08/01/13 87:12 - 88:10, 113:3 - 114:3, 257:23 - 259:13; Holste 07/29/13 51:3 - 53:6, 55:22 - 57:4; Vailhe 6/20/13 182:2 - 185:5.

⁵⁷ B. Todd Heniford 2007 "The benefits of lightweight meshes in Ventral Hernia Repair in Ventral Hernia Repair" Video produced by Ethicon.

⁵⁸ Eth.Mesh.05918776 5/04/04 Email from Schiaparelli, Jill, Strategic Grown Subject: Marlex Experience; Batke 08/01/13 87:12 - 88:10, 113:3 - 114:3, 257:23 - 259:13; Holste 07/29/13 51:3 - 53:6, 55:22 - 57:4; Vailhe 6/20/13 182:2 - 185:5.

⁵⁹ Heniford Video, *supra*, n. 36.

⁶⁰ Id.

In numerous emails, Ethicon employees discussed concerns regarding fibrotic bridging, including concerns about increased risk of contraction of the mesh, nerve entrapment, erosions and chronic pelvic pain if fibrotic bridging occurs.⁶¹ In other emails discussing these concepts, Ethicon's World Wide Marketing Director for General Surgery, Marty Chomiak, states that "... we want to avoid 'bridging', therefore with think large pores are better than small . . ."⁶²

Not only did Ethicon know about the problems with heavy weight, small pore mesh, it also had information and knowledge regarding superior mesh designs to prevent fibrotic bridging and scar plating. Specifically, Ethicon knew that light weight, large pore mesh could decrease the likelihood of foreign body reaction, fibrotic bridging and scar plating.⁶³

Despite having clinical knowledge of the importance of pore size to successful outcomes, and dozens of emails about the importance of pore size, Ethicon's person most knowledgeable about pore size testified that Ethicon does not manufacture its mesh to a specific pore size. Dan Smith testified as follows:

Q: Does Ethicon have a validated test method to determine the pore size of its TVT mesh?

A: We determine the pore size by courses and wales and that is how it's done. So the courses and wale count is a validated test method.

⁶¹ ETH.MESH.04037600 (Innovations in mesh development); ETH.MESH.05920616 (7/20/07 ; Emails from Chomiak, M. to Batke, B., et al. re Defining light weight mesh); ETH.MESH.05585033 (Boris Batke Presentation – Project Edelweis – Ultrapro); ETH.MESH.05446127 (3/13/2006 Emails from Holste, J. to Engel, D., et al. re Mesh and Tissue Contraction in Animal – “Shrinking Meshes?”); ETH.MESH.05475773 (2/09/2007 Boris Batke, Ethicon R&D, Presentation: *The (clinical) argument of lightweight mesh in abdominal surgery*); ETH.MESH.04015102 (3/1/12 Email from Batke, Boris to Mayes, C. re AGES Pelvic Floor Conference-Gala Dinner Invitation); ETH.MESH.04037600 (3/15/12 Boris, B. PowerPoint Presentation, *Innovations in Mesh Development*, Melbourne AGES 2012).

⁶² Eth.Mesh.05920616 7/20/07 Email from Chomiak, Marty Subject: Defining Light Weight Mesh.

⁶³ Batke 08/01/13 87:12 - 88:10, 113:3 - 114:3, 257:23 - 259:13; Holste 07/29/13 51:3 - 53:6, 55:22 - 57:4; Vailhe 6/20/13 182:2 - 185:5; Eth.Mesh.04941016 Lightweight Mesh Developments Dr. Jorge Holste, Senior Research Fellow; Eth.Mesh.01203957 The future of surgical meshes: the industry's perspective Dr. Piet Hinoul November 14-15 2008 Graz, Austria; Eth.Mesh.05916450 Chronic Pain Prevention/future – Bioengineer's point of view Boris Batke Associate Director Research and Development; Eth.mesh.08315779 Clinical Expert Report Prolift +M Dr Piet Hinoul September 25, 2012; Eth.mesh.01752532 Mesh design argumentation issues Juergen Trzewik Ethicon R&D; Eth.mesh.00074499 Gynecare Polift +M Pelvic Floor Repair System Training Presentation.

Q: And I'm talking about pore size. Does Ethicon have a validated test method to determine its pore size for its mesh?

A: The construction of the mesh is -- does not have a pore size requirement.⁶⁴

In fact, Ethicon does not even have a test to measure the pore size of its mesh. Dan Smith testified:

Q. Mr. Smith, does Ethicon have a validated test to describe the pore size of its TVT meshes microns? Yes or no.

A. No...⁶⁵

Despite this information that it did not measure pore size or manufacture its mesh to a specific requirement, Ethicon repeatedly stated in advertising and marketing materials that its mesh was “large pore.” For example, in one brochure, Ethicon promotes the mesh used in the TVT family of products as the “largest pore size” of any of its competitors, listing the size as 1379 um.⁶⁶ However, given that Ethicon has no verified methodology to measure pore size, Ethicon had no scientific basis upon which to base these statements. In fact, in internal documents, Ethicon scientists described Prolene mesh as small pore: “Standard Mesh PROLENE small pores area weight 105 g/m2.”⁶⁷ One Ethicon Engineer measured a mesh and determined that there were two pore sizes in the mesh, a “major” and “minor” pore. “There are two distinct pore sizes in the PROLENE 6 mil mesh (TVT). The major pore is about 1176 um ... The minor pore is about 295 um.”⁶⁸ Certainly, neither of these pores was 1379 um, and the minor pore was substantially smaller.

⁶⁴ Smith Dep. (2-3-14) 729:1 to 729:12

⁶⁵ Smith Dep. (2-3-14) 779:5 to 779:8

⁶⁶ ETH.MESH.00349508 at 9510.

⁶⁷ ETH.MESH.04941016.

⁶⁸ ETH.MESH.00584175 (Ex. T-3583); ETH.MESH.00584179 (Ex. T-3581).

In addition, the pore size of the TVT mesh can change when it is put under stress such as when a sheath is removed or the mesh is tensioned. Dan Smith agreed that these stresses can make an effective pore size smaller than 1 mm.

Q. You would agree, Mr. Smith, that if the measurement across the pores we're looking at here -- let's assume you measure across one of those pores and let's say it's more -- let's say it's 1 millimeter across hypothetically. If a load is put on the mesh and it changes the pore size, that pore could be, after a load is put on it, under 1 millimeter; correct?

A: It's possible depending on the load.⁶⁹

Ethicon engineer Christophe Vaihle testified that “excessive tension on the mesh would lead to the decrease in pore size that can lead to poor tissue integration”⁷⁰

Even though Ethicon has known about the problems related to heavy weight, small pore mesh like the Prolene mesh and the harms and hazards it causes in women’s pelvic tissues, Ethicon has done nothing to change the mesh and continues to promote and sell the product with the same, heavy weight, thick filament “Old Construction 6 mil” mesh that they have been selling since 1974 (Prolene).⁷¹ This is true despite the fact that Ethicon sells large pore, lightweight mesh (which it calls “revolutionary”) for other pelvic floor polypropylene mesh products.⁷² Ethicon believes and tells physicians that the larger pore, lighter weight mesh used for pelvic floor repair is safer and better because it will not cause bridging fibrosis and as much pain for patients.⁷³ Based on my experience, training, and my review of all the scientific literature, I agree that larger pore meshes which are lighter in weight cause less adverse reactions including pain and

⁶⁹ Smith Dep. (2/3/2014) 816:5 to 816:15.

⁷⁰ Vailhe Dep., (6/20/13) 224:10-226:21.

⁷¹ ETH.MESH.03905968

⁷² ETH.MESH.03905968; *see also* Prolift +M CER (“As the mass of a mesh implant is reduced and the pore size is increased, the surface area exposed to the host is reduced, and the foreign body reaction to the implant is reduced.”).

⁷³ ETH.MESH.00074499, ETH.MESH.04941016, ETH.MESH.05479411

erosions for patients.⁷⁴ I agree with Ethicon's scientists, their consultants, and the scientific literature that Ethicon's large pore, light weight mesh designs (i.e. Ultrapro or Prolene Soft) are safer alternatives for permanent implantation in women's pelvises and I believe a larger pore and lighter weight mesh should have been utilized in the TVT.

In summary, for the reasons set forth above, it is my opinion to a reasonable degree of medical certainty that the Prolene polypropylene mesh in the TVT causes fibrotic bridging in the body, resulting in an increased inflammatory response leading to a multitude of injuries, including the possibility of multiple erosions that can occur throughout one's lifetime, chronic and debilitating pelvic pain, recurrence, worsening incontinence, dyspareunia that can be chronic, wound infection, rejection of the mesh, sexual dysfunction, urinary and defecatory dysfunction, vaginal scarring, wound healing problems, injury to ureters, bladder and urethra, pelvic abscess formation, risk of infection, and/or the need for additional surgeries, among others. As a result, the polypropylene in Ethicon's TVT mesh (Prolene) is not suitable for its intended application as a permanent prosthetic implant for stress urinary incontinence in women.

5. *Mesh Contracture/Shrinkage*

Mesh contracture or shrinkage is an event that takes place after the implantation of mesh and relates to the wound healing process that occurs after the surgical trauma of implanting a foreign body made of polypropylene in the sensitive tissues of the vagina and the pelvis. By 1998, polypropylene mesh was known to contract or shrink 30-50%.⁷⁵ Contraction or shrinkage has been shown to draw nerves close to the midurethral sling mesh both in the transobturator

⁷⁴ It should be noted that I have reviewed extensive literature and reports related to histopathology of explanted TVT mesh during the course of my work in this case, including work by Dr. Uwe Klinge, which has shown extensive fibrotic bridging present in virtually all explants of TVT.

⁷⁵ Klinge, U, "Shrinking of Polypropelen Mesh in Vivo: An Experimental Study in Dogs," *Eur J Surg*, 1998 (164:965-969); Jacquetin, B, "Complications of Vaginal Mesh: Our Experience," *Intl Urogyn J*, 2009 (20:893-6); Tunn, R, "Sonomorphological Evaluation of Polypropylene Mesh Implants After Vaginal Mesh Repair in Women with Cystocele or Rectocele," *Ultrasound Obstetrics Gynecol*, 2007 (29:449-452).

application⁷⁶ and for retropubic application.⁷⁷ Furthermore, contraction or shrinkage is closely related to the pore size of the mesh. Small pores lead to fibrotic bridging leading to scar plates, mesh encapsulation and shrinkage or contraction of the mesh, which is a compound and combines with the effect of the normal wound healing process that is already occurring in the tissue.

This phenomenon of shrinkage and its relation to the design of the pores as well as the consequences to the patient were illustrated in an email by Ethicon Scientist Joerg Holste in a March 13, 2006 email discussing a paper he authored entitled “Shrinking Meshes?”⁷⁸ In his email, Dr. Holste states “this was our scientific statement on mesh shrinkage: Basically, small pores, heavy weight meshes induce more fibrotic bridging tissue reaction causing more mesh shrinkage during maturation of the collagenous tissue. See my presentation about biocompatibility.”⁷⁹ In addition, in a presentation by Boris Batke, Associate Director R&D, he states heavier-weight polypropylene mesh results in mesh contraction of 33%.⁸⁰ Ethicon was aware that the mesh in TVT would shrink as well. Specifically, in an email dated November of 2002, related to a discussion of mesh used in a TVT product, it states that Axel Arnaud, one of Ethicon’s medical directors, used 30% shrinkage of the mesh as a “rule of thumb.”⁸¹ At an Ethicon expert meeting in Norderstedt, Germany in 2007, an Ethicon employee presented a

⁷⁶ Corona, R, et al., “Tension-free Vaginal Tapes and Pelvic Nerve Neuropathy,” J Min Invas.Gynecol, 2008 (15:3 262-267); Parnell, BA, et al., “Genitofemoral and Perineal Neuralgia after Transobturator Midurethral Sling,” Obstet Gynecol, 2012 (119:428-431).

⁷⁷ (Heise CP, et al., “Mesh Inguinodynia: A New Clinical Syndrome After Inguinal Herniorrhaphy?” J Am Coll Surg 1998 (187:5 514-8); Voeller, GR, Surg. Technol. Intl. 2003.

⁷⁸ Eth.Mesh 05446127, *supra*, n. 34.

⁷⁹ Id.

⁸⁰ Eth.Mesh 05479717 3/1/11 Boris Batke, Ethicon Associate Director R&D, Presentation: Ethicon Polypropylene Mesh Technology.

⁸¹ Eth.Mesh 03917375

PowerPoint entitled “Factors Related to Mesh Shrinkage” in which all of these issues were clearly laid out.⁸²

Mesh shrinkage was known by Ethicon as early as 1998 in published work by Ethicon’s then consultants, Uwe Klinge and Bernd Klosterhalfen.⁸³ They noted in these early papers that all polypropylene meshes shrink 30-50%. This was restated in later works by W Cobb and his colleagues⁸⁴ – one of which was Dr. Heniford (referenced above). The works of Cobb and Klinge/Klosterhalfen have been referenced in numerous Ethicon documents and thus, Ethicon was well aware of these findings regarding the shrinkage or contraction of polypropylene meshes in vivo. Ethicon was further aware that heavier weight meshes led to greater amounts of contraction.⁸⁵

It is my opinion to a reasonable degree of medical certainty that as a result of work with internal and external experts and consultants in the late 1990s, multiple internal documents and articles, and the scientific literature as a whole, Ethicon was or should have been aware that shrinkage of its Prolene mesh not only could, but would, occur and that this shrinkage could lead to painful complications in women implanted with TVT, such as multiple erosions that can occur throughout one’s lifetime, chronic and debilitating pelvic pain, recurrence, worsening incontinence, dyspareunia that can be chronic, wound infection, rejection of the mesh, sexual dysfunction, urinary and defecatory dysfunction, vaginal scarring, wound healing problems, injury to ureters, bladder and urethra, pelvic abscess formation, risk of infection, and/or the need for additional surgeries, among others. As a result, the polypropylene in Ethicon’s TVT mesh

⁸² Eth.Mesh. 02017152, Nordestadt Expert’s meeting 2007, Eth.Mesh.01782867, “Factors Related to Mesh Shrinking.”

⁸³ Klinge U, Klosterhalfen B, Muller M, Ottinger A, Schumpelick V. Shrinking of Polypropylene Mesh in vivo: An Experimental Study in Dogs. Eur J Surg. 1998; 164; 965-969

⁸⁴ Cobb W, Kercher K, Heniford T. The Argument for Lightweight Polypropylene Mesh in Hernia Repair. Surgical Innovation. 2005; 12(1):T1-T7

⁸⁵ Id.

(Prolene) is not suitable for its intended application as a permanent prosthetic implant for stress urinary incontinence in women,

6. *Fraying, Particle Loss, Sharp Edges, Roping and Curling, Deformation, and Loss of Pore Size*

Since the development of TVT, Ethicon has been aware that Prolene tape frays upon stretching.⁸⁶ In fact, Ethicon designed the TVT mesh, such that, when stress was put on the mesh, particles would separate from the mesh – this was called fraying or linting.⁸⁷ One of Ethicon’s engineers described this as a “defect” that resulted from the method of cutting the mesh: “The mesh frayed is the reverse defect of the mesh features (elasticity of the mesh is one of the commercial arguments to market the TVT)... [T]he root cause of this phenomenon are known: the way to cut the mesh (blade cutting). If we change the way to cut the mesh (ultrasonic cutting or laser cutting) it seems we can limit the mesh frayed defect significantly....”⁸⁸

As early as 2000, Ethicon was aware that particles from TVT Prolene mesh fell into women’s tissues during the sheath removal.⁸⁹ In April 2001, Dr. Alex Wang, “one of the most experienced TVT users in the world,” reported problems with frayed mesh and uneven tape width.⁹⁰ Although the issue was described as “importance: high” requiring prompt attention and action, Ethicon Medical Director, Dr. Martin Weisberg, simply concluded that the deformity in the mesh is undetermined as to whether it has any clinical significance. Dr. Weisberg testified that although he did not actually know whether frayed mesh leading to particle loss would have

⁸⁶ Weisberg 5/31/00 461:7-462:3.

⁸⁷ Weisberg Dep. (5/31/13) 461:7-462:3 (“Q. So engineers within the company knew that fraying of the product was inherent in the design? A. Yes.”).

⁸⁸ ETH.MESH.01813975 at 2 (Ex. 3160/3587).

⁸⁹ Eth.Mesh.01317515 7/12/00 Preventia TVT-2 Risk Analysis Procedure/Tensioning Frayed Mesh/Particle Loss at Eth.Mesh.01317523.

⁹⁰ Eth.Mesh.03905472 6/4/01 Emails from Wang, Dr. Alex Subject TVT Recommendation for Ethicon Study of Fraying/Particle Loss.

clinical implications, he does not recall whether he or anyone else at Ethicon studied the issue.⁹¹

Just a few months later, however, Ethicon received a complaint by an experienced surgeon regarding a patient who experienced vaginal wall erosion following a TVT procedure which was first noted by her husband during intercourse. According to the surgeon, “the tape appeared frayed and tiny fibers were protruding through the vaginal wall.”⁹²

In November 2003, Dr. Weisberg reported that there had been a total of 58 complaints of fraying with TVT since introduction of the device in 2000. He observed that the following occurs when the mesh frays: “[T]he mesh elongates in places; the mesh narrows in places; and small particles of Prolene might break off ... and that [s]tretching of the mesh increases the probability of fraying.”⁹³ Once again, however, Dr. Weisberg concluded that “since fraying does not affect the safety and efficacy of the TVT device, it has been determined not to pursue any corrective actions at this time.”⁹⁴ Dr. Weisberg confirmed during his deposition that no corrective action was taken and, although he did not know whether Prolene particles could elicit a chronic foreign body response, he does not recall whether he or anyone else at Ethicon investigated the issue.⁹⁵

In 2004, Ethicon continued to receive complaints from surgeons about fraying and “brittle” mesh and particles falling into the operating field.⁹⁶ One of the company’s “most urgent customers,” Swiss surgeon Dr. J. Eberhard, wrote the following: “Already at the operation it is embarrassing to see how the tape is crumbling. But it gets worse if there is stretch on the tape.

⁹¹ Weisberg 5/31/13 469:23-470:16.

⁹² Eth.Mesh.02621559 at Eth.Mesh.02622276 Ethicon Issue Report TVT Retropubic 2001 Open Date Between 01-Jan-2001 and 31-Dec-2001.

⁹³ Eth.Mesh.00541379 11/18/03 Memo from Weisberg, Dr. Martin Subject: Mesh Fraying for TVT Devices Inadequate Testing.

⁹⁴ Ibid.

⁹⁵ Weisberg 5/31/13 469:23-470:16.

⁹⁶ Eth.Mesh.00863391 at Eth.Mesh.00863392 2/27/04 Emails from Smith, Dan Subject 2 TVT Complaints Concerning Allegedly Brittle Mesh.

... I can't understand, that no one will solve that problem for such a long time. As the latest, as the tape has becoming blue, everyone has realized, that the quality of the tape is terrible.”⁹⁷ Dan Smith, a Lead Engineer on TVT products, lamented the particle loss that was revealed when the mesh was dyed blue: “This is not going away anytime soon and competition will have a field day, major damage control offensive needs to start to educate reps and surgeons UPFRONT that they will see BLUE shit and it is OK.”⁹⁸ Indeed, in November of 2004, one of the “top 3 complaints” included “mesh frayed.”⁹⁹ Once again, however, Ethicon decided to take no corrective action.¹⁰⁰ Instead, sales representatives were instructed to reassure their doctors that, “Prolene is proven to be inert,” the “particles will not cause any problem,” and to “be proactive” because “the competition will try to target this!”¹⁰¹ Physicians were told the particles are “non-reactive” and that fraying does not affect the safety or efficacy of the device.¹⁰² In fact, it has consistently been Ethicon's position that frayed mesh and resulting particle loss as well as roping, curling and deformation of the mesh do not create a safety risk and have no clinical significance.¹⁰³ However, as noted above, Ethicon never tested whether the particles would have a negative clinical impact on women. An independent investigator, Dr. Pariente, did and published a study that concluded that “the very high particle shedding for both Sparc (AMS) and TVT (Ethicon) may be of significant long term clinical concern in some quarters.”¹⁰⁴ Dr.

⁹⁷ Eth.Mesh.02180833 11/12/04 Letter from Prof. Dr. Eberhard (translated); Eth.Mesh.02180828 11/12/04 Telefax from Sibyll, Basso to Menneret, David re Prof. Dr. Eberhard.

⁹⁸ ETH.MESH.00863391.

⁹⁹ ETH.MESH.01813975 (Ex. T-3160 / T-3587).

¹⁰⁰ Eth.Mesh.02180826 11/12/04 Email from Menneret, David to Smith, Dan and others Subject Mesh Fraying: Dr. Eberhard Letter.

¹⁰¹ Eth.Mesh.00865322 3/2/04 Email from Bell, Steve, Ethicon Marketing Director Europe to Sales & Marketing Team Subject: Reminder on Blue Mesh – Frayed Mesh/Particle Loss.

¹⁰² Eth.Mesh.03535750 10/12/2005, Hunsicker, MSN, CRNP, Kimberly, Ethicon Clinical Operations Regional Manager, Presentation: Investigator Initiated Study Process – Inadequate Testing.

¹⁰³ Eth.Mesh.00541379, *supra*, n. 58; Eth.Mesh.00858252 2004 Memo from London Brown, Allison, to Smith, Dan Subject Mechanical Cut v. Laser Cut Mesh Rationale; *See also*, [Eth.mesh.03924557](#).

¹⁰⁴ Eth.Mesh.01221055 Pariente, J-L, “An independent biomechanical evaluation of commercially available suburethral slings,” Issues in Women's Health, 2003.

Pariente also noted that during surgical use, these particles are released into soft tissue and it is not possible to know where they go.¹⁰⁵ Although Ethicon claims that its own internal testing shows approximately 1% particle loss with TVT,¹⁰⁶ Dr. Pariente's study demonstrated TVT particle loss as high as 8.5% - 10 times higher than most of its competitors.¹⁰⁷ In addition, in a "Pelvic Floor Repair-Surgeon's Feed-back on Mesh Concept" summary, a number of physicians informed Ethicon that "small particles are released that migrate through the vaginal wall causing pain" and that "it is of utmost importance that the mesh is cuttable and that it does not fray nor release particles after cutting. The small particles migrate and cause pain during intercourse."¹⁰⁸ Surgeons also reported to Ethicon that upon handling Ethicon's Prolene mesh, there was an "immediate release" of "a lot of particles" from the mesh. One surgeon stated he would have less anxiety using a mesh that did not fray like Ethicon's Prolene mesh.¹⁰⁹ Even Ethicon's own paid consultant surgeons expressed concern on behalf of their surgeon colleagues over leaving frayed mesh particles in their patients.¹¹⁰ These same surgeons told Ethicon that a mesh without particle loss "was definitely a needed improvement."¹¹¹

Although Ethicon knew that particles of the TVT mesh were falling into women's delicate vaginal tissues since the launch of the TVT, and knew that these particles could cause pain and discomfort for women since at least 2000 or 2001, it wasn't until late 2006 or early 2007 that Ethicon finally sold a laser cut mesh that reduced the amount of "non-functioning"

¹⁰⁵ *Ibid.*

¹⁰⁶ Eth.Mesh.000585802, or 6/12/06 Kammerer 00585842; Eth.Mesh.00585823 06/27/06 Email from Kammerer, Gene to Volpe, Clifford, Subject GY: ***URGENT*** French STANDARD ON TVT & MESHES (COMMENTS REQUIRED)

¹⁰⁷ Eth.Mesh.01221055, *supra*, n. 67; Eth.Mesh.00585842 6/12/06 Email from Kammerer, Gene to Rha, Sunny Subject TVT LCM – Particle Loss (Reimbursement Submission); Eth.Mesh.01219629 5/09/06 Email from Flatow, Jacqueline to Kammerer, Gene Subject: Re: Particle loss on TVT; Eth.Mesh.01221024 Email 5/04/06 from Kammerer, Gene to Fourneir, Herve and Arnaud, Axel Subject: New Standards for Urethral Slings; Eth.Mesh.00585823, *ibid.*

¹⁰⁸ Eth.Mesh.05644163-05644171

¹⁰⁹ *Id.*

¹¹⁰ ETH.MESH.01809082.

¹¹¹ *Id.*

material that would fall off the TVT mesh. Ethicon's April 2006 Clinical Expert Report for Laser Cut Mesh stated that there was a decrease in particle loss with laser cut mesh and this "decrease would lead to less non-functioning material left in the tissues."¹¹² It cannot be disputed that the greater the nonfunctioning material left in a patient's tissue, the greater the surface area of polypropylene the patient is exposed to, and the greater the inflammatory responses and the greater the foreign body response. As discussed above, the long term consequences of this chronic foreign body reaction and inflammatory response can be, among other things, chronic pain, lifelong risk of erosions, dyspareunia and failure of the device. If the individual flakes work their way through the vaginal mucosa, this can lead to dyspareunia and/or painful intercourse for the partner as noted in the complaints received by Ethicon back in 2001 (referenced above). The larger the surface area, the greater the risk associated with vaginal mesh. Finally, detached flakes of polypropylene may migrate into the vasculature or lymphatics and cause problems remote from the pelvis. For these reasons, Ethicon should have used a mesh without the fraying and particle loss defect when selling its TVT for permanent implant in a woman's vaginal tissues.

In addition to the fraying and particle loss defects, the mechanically cut mesh used in the TVT has also been shown to rope, curl and easily deform when placed under very little tension. Ethicon also learned about these defects very shortly after it started marketing the TVT mesh. Not surprisingly, Ethicon's response to learning about these defects was similar to its response to learning about the fraying and particle loss defects – that is, Ethicon claimed that these problems

¹¹² ETH.MESH.00167104 at 7109.

with the mesh did not have any clinical significance despite the fact that surgeons were stating otherwise.¹¹³

The degree of roping, curling, degradation and deformation of the TVT mesh when under tension is well established and, further, is very substantial. The findings of outside researcher Pamela A. Moalli, M.D., Ph.D., a urogynecologist and professor at the University of Pittsburg, showed “that although very little force [was] applied” the TVT Prolene mesh was subject to “permanent deformation.”¹¹⁴ Dr. Moalli and her bioengineering colleagues at the University of Pittsburg engineering school performed an in-depth analysis in which the biomechanical properties of six mid-urethral slings, including TVT Prolene mesh, were tested. The research included different tests which analyzed the load or stress placed on the slings. The researchers “were primarily interested in comparing the relative resistance of the slings to deformation (stiffness), the degree to which the slings deformed after application of a load (relative elongation), and their overall biomechanical behavior.” *Id.* The testing revealed “easy permanent deformability” of the TVT Prolene mesh “*that is observed clinically during placement.*”¹¹⁵ The authors wrote, “one of the primary problems in using the TVT is that as a result of its low stiffness, the mesh easily deforms when tensioning under the urethra. Specifically, pulling the sling gently results in thinning of the mesh (permanent deformation) and fraying at the tanged edges.”¹¹⁶ The researchers found that other slings were more resistant to deformation than TVT Prolene mesh.

In 2006, Ethicon did its own analysis of whether the mechanically cut TVT mesh would rope, curl, deform and lose particles under tension that could exist in the real world.

¹¹³ Eth.Mesh 00440005; Eth.Mesh 00302390 TVT-Base & TVT-O Review for Laser Cut Mesh (LCM) Risk Analysis


¹¹⁴ Eth.mesh.00294195 (Moalli et al, Tensile properties of five commonly used mid-urethral slings relative to the TVT, International Urogynecology Journal, 2007).

¹¹⁵ *Id.* (emphasis added).

¹¹⁶ *Id.*

Specifically, an Ethicon Engineer responsible for the laser cut mesh project, Gene Kammerer, made a presentation that clearly showed each of these defects in the mechanically cut mesh.

These photos clearly show particle loss, fraying, degradation, roping and deformation when the mechanical cut mesh was stretched and compared to TVT Laser Cut.¹¹⁷



As noted these photos show mesh after 50% elongation. I have read depositions of Ethicon personnel claiming this is not a realistic elongation seen with mesh. However, Ethicon's engineer who performed the study (and took the photos), Gene Kammerer, explained as follows why he chose 50% elongation for the study and explained that he had experienced 50% elongation himself during testing of the mesh:

The link between the elongation percent, not force, and the integrity of the mesh is this. During the operative procedure as the surgeon removes the protective sheath from the mesh, the mesh stretches or elongates. It is my

¹¹⁷ ETH.MESH.08334245.

experience, after viewing many surgical procedures and performing numerous procedures on cadavers myself, that the mesh stretches approximately 50% at the maximum. There is also additional stretching that occurs if the surgeon elects to do an adjustment on the position of the mesh under the urethra. It is these two occurrences which produce the majority of the particle loss and loss of the integrity of the construction of the mesh.¹¹⁸

Again, Ethicon claimed that these problems with the mesh did not have any clinical significance despite the fact that surgeons were complaining.¹¹⁹ However, Ethicon's own internal documents demonstrate that this is not true. According to Ethicon's Failure Modes Analysis and other internal documents, the loss of pore size due to mesh narrowing or deformation can lead to urinary retention or erosion.¹²⁰ Ethicon's own dFMEA shows that the hazards of curling/roping, frayed edges and inadequate pore size of mesh can lead to the harms of erosion, recurrence, and pain.¹²¹ When discussing the dFMEA for Laser Cut Mesh, Former Medical Director, David Robinson, agreed that pore size of both the Laser Cut and Mechanically Cut mesh "[c]ould reduce, the tissue might not encapsulate . . . the tissue might not grow through the mesh. It can become encapsulated and then it could cause . . . a rejection of the mesh."¹²² And, Dr. Robinson testified that a rejection of the mesh can lead to an erosion.¹²³ These changes in the mesh may lead to erosion or pain for women with the deformed mesh implanted in their bodies. Further, according to Ethicon, this curling, roping, or narrowing of the mesh may also cause urinary retention in addition to erosion and pain.¹²⁴ Engineers at Ethicon knew that the TVT mesh could cause more urinary retention than some of its other meshes because the mesh

¹¹⁸ ETH.MESH.00584811.

¹¹⁹ ETH.MESH 00440005; ETH.MESH 00302390 (TVT-Base & TVT-O Review for Laser Cut Mesh (LCM) Risk Analysis).

¹²⁰ Eth.Mesh.01218019;

¹²¹ Eth.Mesh.01218019.

¹²² Robinson 9/11/13 1070:23-1072:25.

¹²³ Ibid.

¹²⁴ Robinson Dep. (9/11/13) 1079:3-4-1081; 1081:9-13; 1083:8-18; ETH.MESH.01218019, ETH.MESH 01822361.

will “curl and rope which reduces the surface area of the mesh under the urethra and therefore, increases the pressure in a localized point.”¹²⁵

In fact, I have witnessed the same type of roping and narrowing of the TVT mesh when I placed them myself and can see the deformed and roped mesh when I remove them. This localized pressure under the urethra leads to complications like, among others, urinary retention, chronic pain, dyspareunia and erosions. In addition, I have reviewed Ethicon TVT training videos that show the exact problem discussed about related to deformation and roping of the “tape” under the urethra.¹²⁶ Finally, according to Ethicon’s Dan Lamont, it chose to continue to sell “mechanically cut mesh despite knowing that it had the potential for degradation, particles floating around in women’s bodies, stretching, and roping . . .”¹²⁷ Lamont admitted that the fraying of the mesh was a “defect” of the mesh.¹²⁸ These defects made the TVT mechanically cut mesh unsuitable for the permanent implantation in women’s vaginal tissues.

Another defect of mechanically cut mesh is its sharp edges as shown on this photo:¹²⁹



¹²⁵ Eth.Mesh.01822361 Email from Dan Smith re TVT Secur.

¹²⁶ Eth.Mesh.PM.000004 TVT Retropubic Implantation Video.

¹²⁷ Lamont 9/11/13 30:18-24

¹²⁸ Lamont 9-11-13, 15:16-16:10

¹²⁹ ETH.MESH.09656795.

While Ethicon states that these sharp edges are part of the intended “velcro” effect of mesh, it was a feature about which Ethicon received multiple complaints tied to injuries and erosions. For example, during one market research test with physicians, it was reported:

The surgeon felt that the MCM strips was elastic but with "hairs" on the edges and that it scratched with abrasive texture scraping (like the Scotch - Brite™ pads), furthermore a lot of particles were released and a rope/string effect could occur if an excessive force was applied.¹³⁰

When one agency recognized a spike in erosions, it inquired whether this was the result of “the cut ends of the tape appear to be sharper and more likely to cut tissue.”¹³¹ A sentiment shared by some physicians and reported to Ethicon:

Basically, he thinks that erosions due to the TVT mesh are underestimated in reports. The reason is that in order to recognize them, a very careful vaginal examination is needed. Most of the time, a "hidden" erosion is asymptomatic and neither the patients nor their sexual partner if any complain. But it might happen that a patient may complain. **He believes that erosion are due to the sharp edges of the mesh.** He wanted to suggest that we add to the mesh edges a kind of seam that would help preventing erosion. (emphasis added)¹³²

Dr. Axel Arnaud responded that Ethicon did not want to modify its mesh (even if it caused erosions) because Ethicon did not want to lose the marketing edge of using the Ulmsten/Nilsson data. He wrote:

I also indicated that we want to be very careful with any modifications of our tape since a change in the mesh would obsolete all the long term clinical results we have about the procedure.¹³³

Ethicon also had multiple reports of frayed, sharp edges of the TVT mesh protruding through women’s vaginal tissues.¹³⁴ I have personally touched the edges of the mechanically cut TVT mesh and have found the edges to be rough and sharp. These rough edges can irritate

¹³⁰ ETH.MESH.06696589

¹³¹ ETH.MESH.00330760

¹³² ETH.MESH.03911107 (Axel Arnaud reporting his interview with Professor Hausler).

¹³³ *Id.*

¹³⁴ ETH.MESH.02620681, ETH.MESH.02625246, ETH.MESH.02654027, ETH.MESH.02653319, ETH.MESH.02653001, ETH.MESH.03715978

tissue, cause increased inflammation and foreign body response, and cause or contribute to cause chronic pain, dyspareunia and erosions. This defect is another reason that the TVT mechanically cut mesh is not a suitable mesh for permanent implantation in women's vaginal tissues.

Ultimately, the market pressure on Ethicon to develop a laser cut mesh without particle loss, roping, curling, deformation, and sharp edges became too much. Paula Evans, Gynecare European Marketing Manager, described the situation as: "France is in a recovery mode, Germany is hemorrhaging business ... Without laser cut, there is the real risk that more business will be lost." (sic).¹³⁵ Hence, the laser cut mesh project went forward to address the chronic problems with particle loss, fraying, roping, curling, sharp edges and deformation encountered by patients with mechanically cut mesh.¹³⁶ In late 2006, Ethicon developed laser cut mesh. Astonishingly, as part of the design verification activities, Ethicon initially decided that particle loss, elongation curve and flexural rigidity data would *not* be required because they were not "critical to quality."¹³⁷ In fact, this news was celebrated as "!!!!GREAT NEWS FOR TVT LASER CUT MESH!!!!" and "less work for all of us."¹³⁸

During the early development of the laser cut mesh, Ethicon acknowledged that mechanical cut mesh and laser cut mesh were two separate mesh products and to imply otherwise would be misleading. In anticipation of launching the new laser cut mesh, though, a dilemma developed regarding how to market the new mesh to doctors. In December 2005, Kevin Mahar described the marketing strategy "...KEEP selling regular TVT (the 'Colonel's Original Recipe') to those customers that want/love it...and KEEP going forward with 8 years of data, etc with the original recipe ... We do not mislead them that this is the same

¹³⁵ ETH.MESH.04985249.

¹³⁶ ETH.MESH.00301741 (11/21/05 Email from Lamont, D. re !!!!GREAT NEWS FOR TVT LASER CUT MESH!!!! –Frayed mesh/particle loss); ETH.MESH.00394544 (2/01/06 Global Regulatory Strategy – GYNECARE TVT – Laser Cutting Project); Weisberg Dep. (5/31/13) 487:13-488:7. v

¹³⁷ Eth.MeshIbid; Weisberg 5/31/13 490:15-491:17.

¹³⁸ Eth.MeshI ibid.

product....”¹³⁹ However, because Ethicon wanted to continue to claim the marketing benefit of the Ulmsten/Nilsson data,¹⁴⁰ Ethicon decided it could not market the laser cut mesh as distinctly different or it risked losing the ability to claim the application of the Ulmsten study data to the new mesh. This was described as a way to protect the “clinical heritage” of the mesh:

Marketing Need: Keep clinical heritage intact.... In order to continue to claim the use of 7-year data and all clinical studies, the MCM and LCM needed to show similar properties with the physical properties being used as a proxy for the clinical needs.¹⁴¹

Ultimately, Ethicon did end up telling doctors that the mechanically cut mesh and the laser cut mesh are essentially the same, a decision that has kept doctors in the dark about the defects inherent in the mechanically cut TVT mesh, and has led to continued harms and hazards to women. In my opinion as a physician, Ethicon’s decision to continue to market and sell the mechanically cut mesh, with all of its defects, and the decision to market the new improved laser cut mesh as virtually the same as the old construction TVT mesh, was clearly a decision by Ethicon to put profits before patient safety.

In summary, for the reasons set forth above, it is my opinion to a reasonable degree of medical certainty that the Prolene polypropylene mesh in the TVT has several characteristics that make it improper for use in a women’s vaginal canal including fraying, particle loss, roping, curling, deformation, sharp edges and loss of pore size. These unwanted characteristics can lead to, among other things, an increased inflammatory response and/or increased pressure on the urethra (roping or curling) or loss of pore size (roping or curling), and can lead to a multitude of injuries, including such as multiple erosions that can occur throughout one’s lifetime, chronic and debilitating pelvic pain, recurrence, worsening incontinence, chronic dyspareunia, wound

¹³⁹ ETH.MESH.00687819 (Ex. T-3164).

¹⁴⁰ ETH.MESH.03911107; Isenberg 11/6/13 421:8-19

¹⁴¹ ETH.MESH.00858252; *see also* ETH.MESH.00526473; ETH.MESH.02248778 (Kammerer PPT); Hellhammer Dep. (9/11/13) 120-121.

infection, rejection of the mesh, sexual dysfunction, urinary and defecatory dysfunction, vaginal scarring, wound healing problems, injury to ureters, pelvic abscess formation, risk of infection, and/or the need for additional surgeries, among others. As a result, the polypropylene in Ethicon's TVT mesh (Prolene) is not suitable for its intended application as a permanent prosthetic implant for stress urinary incontinence in women.

Moreover, Ethicon did not inform physicians and patients that its mesh was susceptible to these physical deformations that could lead to painful erosions, recurrent, late infections and the need for mesh removal. Nor did Ethicon inform physicians that laser cut mesh had materially different mechanical properties than mechanically cut mesh. By failing to do so, Ethicon did not adequately warn physicians about these important risks, nor by extension, provide surgeons with an opportunity to discuss these risks with their patients.

B. Ethicon knew that the old construction mesh (Prolene) was not appropriate for use in its TVT device as early as 1998, but failed to modify/change the mesh to a larger pore, lighter weight mesh that would not deform, fray, lose particles, rope, curl, degrade, cause excessive foreign body reactions, and cause excessive shrinkage/contraction because of its economic interest in maintaining its competitive advantage in the MUS market and, therefore, Ethicon put profits before patient safety;

As stated above, Ethicon knew from the time it launched the TVT with the mechanically cut mesh that it was defective in multiple respects. This is true because the TVT Prolene mesh was known to be made from heavyweight 6 mil fiber and a construction that allowed for mesh curling, roping, fraying, zipping, particle loss, and sharp edges. In fact, beginning in 1998, Ethicon had already established a "mesh improvement project" in order to improve the mesh. Despite the fact that the project yielded an improved mesh, Ethicon never incorporated those improvements into the TVT.

As early as May of 1997, Ethicon knew that the Prolene mesh was not ideal for use in vaginal tissues.¹⁴² In fact, Ethicon knew of a case at that time where a patient had been treated with Prolene mesh, which protruded through the vagina, requiring excision of the mesh. Ethicon knew that the ideal mesh for use in the vagina should not have any fraying or spiky edges, needed to have large enough pores to encourage in-growth, and should have a low mass density to minimize foreign body reaction.¹⁴³ Ethicon then embarked on a project to improve the Prolene mesh used in the TVT product and Ethicon's hernia products. Among the characteristics they sought to improve were the product curling, zipping and unraveling of the mesh after cutting, and crumbling of the mesh.¹⁴⁴ Ethicon noted that if the Prolene mesh was pulled in one direction, the mesh would curl up into a tube, and the mesh would remain in a rolled condition even after the force of the pulling was no longer on the mesh.

¹⁴² ETH.MESH.12006257

¹⁴³ ETH.MESH.12006257

¹⁴⁴ ETH.MESH.09264945




Figure 1 – Control mesh sample before and after the application of the force. A clear picture of mesh curling results.

Ethicon also referred to the original construction 6 mil Prolene mesh as a mesh that was known for its “bad” curling quality.¹⁴⁵ Ethicon ultimately changed the flat Prolene mesh used for hernia repair to address these issues, making changes to the construction of the mesh to address the bad curling quality of the mesh, and at the same time, changing to a lightweight, 5 mil mesh construction.¹⁴⁶ The change in the mesh construction also made the mesh less likely to fray and lose particles.¹⁴⁷ Despite Ethicon’s original intent to incorporate the new construction material which was lighter weight and had improved resistance to curling, fraying, and particle


¹⁴⁵ HMESH_ETH_00946834, ETH.MESH.02182844

¹⁴⁶ HMESH_ETH_00782152

¹⁴⁷ HMESH_ETH_02008684

loss,¹⁴⁸ Ethicon continued and still continues to use the original, old, old heavyweight 6 mil construction mesh for the TVT products.¹⁴⁹

The flaw in the construction of the TVT heavyweight Prolene mesh which allows it to curl into a tube after tensioning or pulling on the mesh and not return to its original shape, combined with the heavyweight and small pore nature of the mesh, causes the mesh to fold up and become hard post-implantation. Ethicon continued to be aware of this continuing defect in the mesh well after the Prolene mesh improvement project was completed and the company changed the construction of its Prolene hernia mesh.¹⁵⁰ Ethicon was also aware that lightweight materials were less likely to fold up post implantation and integrated better with surrounding tissues,¹⁵¹ but continued to use the heavier 6 mil fibers. The lightweight materials were also much better at resisting crumpling and less likely to have sharp edges during tissue integration.¹⁵²



Traditional polypropylene mesh. 90 days post-implantation. Fold development (in-vivo study)



Lightweight VYPRO II mesh. 90 days post-implantation. Fold-free incorporation (in-vivo study)

¹⁴⁸ ETH.MESH.09264884

¹⁴⁹ HMESH_ETH_02030355, ETH.MESH.09275875

¹⁵⁰ ETH.MESH.05918776

¹⁵¹ ETH.MESH.05446129

¹⁵² Ethicon Tissue Reinforcement Solutions, 8/21/2004

Ethicon continued to have problems with mesh quality in the TVT mesh after the Prolene mesh improvement project was complete, but never incorporated those changes into the TVT mesh. After the improved construction 5 mil Prolene mesh replaced the 6 mil mesh Prolene mesh for flat hernia repair, Ethicon noted continuing problems with the Prolene mesh in the TVT, noting inconsistent tape width,¹⁵³ and fraying and particle loss from the TVT mesh.¹⁵⁴ Doctors reported to Ethicon that the quality of the mesh was terrible, and that particles were falling off the mesh, which was worse when the mesh was elongated.¹⁵⁵

Ethicon did not change the Prolene mesh in its TVT device despite having better and safer options available for economic reasons. Ethicon believed that continued use of the TVT mesh gave the company an economic and competitive advantage in marketing the product because they could continue to use the existing clinical data on the product to market the device, while if the mesh was changed, the existing clinical data would be obsolete.¹⁵⁶ Dr. Brigitte Hellhammer testified that despite having incorporated the use of the lightweight, large pore Ultrapro mesh in vaginal tissues for the treatment of pelvic organ prolapse, the Ultrapro was never used by Ethicon in a device used for the treatment of stress urinary incontinence largely because the company wanted to continue to rely on the Ulmsten/Nilsson series of studies on 130 patients performed with the TVT device.¹⁵⁷ Dr. Arnaud also confirmed that the company did not want to change anything with the mesh because of the exiting clinical data on the product.¹⁵⁸ It is my opinion to a reasonable degree of medical certainty that Ethicon was negligent in failing to correct the defects in the TVT mesh as the company had knowledge of the defects and failed to

¹⁵³ ETH.MESH.12002601

¹⁵⁴ ETH.MESH.00863405,

¹⁵⁵ ETH.MESH.02180833

¹⁵⁶ ETH.MESH.03911107

¹⁵⁷ Deposition of Brigitte Hellhammer, MD, September 11, 2013

¹⁵⁸ Deposition of Axel Arnaud, July 19, 2013 36:15-37:3

correct the defects with products and solutions that were already available to the company because it put its economic interests above the safety of patients.

C. Ethicon's Warnings and Disclosures of Adverse Events in its TVT IFU Have Been Inadequate Based on the Adverse Reactions and Risks Associated with the TVT That Have Been Known to Ethicon from the time the TVT was first sold and marketed.

1. Ethicon failed to include multiple adverse reactions and risks associated with the TVT in the IFU.

The purpose of the IFU is for a medical device manufacturer to provide physicians with the information necessary for them to make decisions regarding the medical device for a particular patient. In addition, the IFU should disclose adverse reactions and risks known to the medical device manufacturer to the physician so that the risks can be relayed to the patient and an informed decision regarding the use of the product can be reached. Throughout my education, training, surgical and clinical practice, I have reviewed numerous IFUs for a variety of products, including mesh products, in order to understand the proper way to use the device and to gain knowledge about the complications and adverse events associated with a device. I have extensive clinical experience with IFUs and instructing patients about the adverse events/risks contained in the IFU. Similar to Ethicon's Medical Directors, Dr. Martin Weisberg and Dr. David Robinson, I have gained expertise in IFUs through my extensive clinical experience reviewing IFUs, consenting patients regarding IFUs, including Ethicon's own pelvic mesh products like the TVT and Prolift.

Catherine Beath, Ethicon's former Vice President of Quality Assurance and Regulatory Affairs, testified that "physicians should be made aware of all the significant safety risks associated with the product in the IFU."¹⁵⁹ And, "a reasonably prudent medical device company

¹⁵⁹ Beath 7/12/13 592:7-11.

would continually update the label consistent with developing data and information that becomes known to the company” when it is appropriate.¹⁶⁰ Similarly, former Medical Director, Dr. David Robinson, testified that the warnings and adverse event section of the IFU should include all significant risks and complications related to the procedure and the mesh.¹⁶¹ According to Dr. Robinson, a device manufacturer must include this information because you want to make sure the doctors have all the information they need to adequately inform patients who are deciding to use the product.¹⁶² According to Ethicon Medical Director, Dr. Martin Weisberg, the goal of the IFU is to communicate the most important safety risks attributable to the TVT device and an IFU should never exclude known hazards or complications.¹⁶³ Dr. Weisberg also confirmed that an IFU should not knowingly underestimate the risks of using the product and, further, that if an IFU¹⁶⁴ excludes known complications or understates risks, it “fails in one of its principal purposes.”¹⁶⁵ Finally, Peter Cecchini, a 43 year Ethicon employee and Regulatory Fellow and the person responsible for the TVT 510K, testified that the “regulatory standard for the IFU is the known risks are supposed to be included in the adverse reactions.”¹⁶⁶ Mr. Cecchini testified that he relies on medical affairs to make sure he knows the known risks so they can be included in the IFU.¹⁶⁷

From September of 2000 to the present day, there have been six versions of the Ethicon TVT IFU. These include the following versions: September 8, 2000, December 22, 2003, February 11, 2005, April 7, 2006, October 13, 2008 and November 29, 2010. A chart showing

¹⁶⁰ Beath, 7/11/13, 198: 8-13

¹⁶¹ Robinson, 9/11/13, 1046:9-13

¹⁶² Robinson, 9/11/13, 1046:23-1047:13

¹⁶³ Weisberg 8/9/13, 959:19-960:15

¹⁶⁴ *Id.* at 960:13-16.

¹⁶⁵ *Id.* at 961:10-17.

¹⁶⁶ Cecchini, 10/22/12, 65:5-12

¹⁶⁷ Cecchini, 10/22/12, 65:18-24

the Adverse Reactions/Risks section for each version of the TVT Instructions for Use is set forth below.

Product	Production Prefix	Start Bates	End Bates	First Use Date	Last Use Date	Adverse Reactions / Risks
<i>TVT</i>	ETH.MES H.	52253 54	52253 85	09/08/00	11/26/03	<ul style="list-style-type: none"> * Punctures or lacerations of vessels, nerves, bladder or bowel may occur during needle passage and may require surgical repair. * Transitory local irritation at the wound site and a transitory foreign body response may occur. This response could result in extrusion, erosion, fistula formation and inflammation. * As with all foreign bodies, PROLENE Mesh may potentiate an existing infection. The plastic sheath initially covering the PROLENE Mesh is designed to minimize the risk of contamination. * Over correction, i.e., too much tension applied to the tape may cause temporary or permanent lower urinary tract obstruction.
<i>TVT</i>	ETH.MES H.	23403 06	23403 69	12/22/03	02/11/05	<ul style="list-style-type: none"> * Punctures or lacerations of vessels, nerves, bladder or bowel may occur during needle passage and may require surgical repair. * Transitory local irritation at the wound site and a transitory foreign body response may occur. This response could result in extrusion, erosion,

						<p>fistula formation and inflammation.</p> <p>* As with all foreign bodies, PROLENE Mesh may potentiate an existing infection. The plastic sheath initially covering the PROLENE Mesh is designed to minimize the risk of contamination.</p> <p>* Over correction, i.e., too much tension applied to the tape may cause temporary or permanent lower urinary tract obstruction.</p>
TVT	ETH.MES H.	23404 71	23405 03	02/11/05	04/07/06	<p>* Punctures or lacerations of vessels, nerves, bladder or bowel may occur during needle passage and may require surgical repair.</p> <p>* Transitory local irritation at the wound site and a transitory foreign body response may occur. This response could result in extrusion, erosion, fistula formation and inflammation.</p> <p>* As with all foreign bodies, PROLENE Mesh may potentiate an existing infection. The plastic sheath initially covering the PROLENE Mesh is designed to minimize the risk of contamination.</p> <p>* Over correction, i.e., too much tension applied to the tape may cause temporary or permanent lower urinary tract obstruction.</p>

TVT	ETH.MES H.	52226 73	522270 4	4/07/06	10/07/08	<p>* Punctures or lacerations of vessels, nerves, bladder or bowel may occur during needle passage and may require surgical repair.</p> <p>* Transitory local irritation at the wound site and a transitory foreign body response may occur. This response could result in extrusion, erosion, fistula formation and inflammation.</p> <p>* As with all foreign bodies, PROLENE Mesh may potentiate an existing infection. The plastic sheath initially covering the PROLENE Mesh is designed to minimize the risk of contamination.</p> <p>* Over correction, i.e., too much tension applied to the tape may cause temporary or permanent lower urinary tract obstruction.</p>
TVT	ETH.MES H.	23405 04	234056 7	10/13/08	11/22/10	<p>* Punctures or lacerations of vessels, nerves, bladder or bowel may occur during needle passage and may require surgical repair.</p> <p>* Transitory local irritation at the wound site and a transitory foreign body response may occur. This response could result in extrusion, erosion, fistula formation and inflammation.</p> <p>* As with all foreign bodies, PROLENE Mesh may potentiate an existing infection. The plastic sheath initially</p>

						<p>covering the PROLENE Mesh is designed to minimize the risk of contamination.</p> <p>* Over correction, i.e., too much tension applied to the tape may cause temporary or permanent lower urinary tract obstruction.</p>
TVT	ETH.MES H.	34278 78	342794 5	11/29/10	To Present Day	<p>* Punctures or lacerations of vessels, nerves, bladder or bowel may occur during needle passage and may require surgical repair.</p> <p>* Transitory local irritation at the wound site and a transitory foreign body response may occur. This response could result in extrusion, erosion, fistula formation and inflammation.</p> <p>* As with all foreign bodies, PROLENE Mesh may potentiate an existing infection. The plastic sheath initially covering the PROLENE Mesh is designed to minimize the risk of contamination.</p> <p>* Over correction, i.e., too much tension applied to the tape may cause temporary or permanent lower urinary tract obstruction.</p>

In all six versions of the TVT IFU from September of 2000 to present day, the Adverse Reactions/Risks section has remained exactly the same. It reads as follows:

ADVERSE REACTIONS

- * Punctures or lacerations of vessels, nerves, bladder or bowel may occur during needle passage and may require surgical repair.
- * Transitory local irritation at the wound site and a transitory foreign body response may occur. This response could result in extrusion, erosion, fistula formation and inflammation.
- * As with all foreign bodies, PROLENE Mesh may potentiate an existing infection. The plastic sheath initially covering the PROLENE Mesh is designed to minimize the risk of contamination.
- * Over correction, i.e., too much tension applied to the tape may cause temporary or permanent lower urinary tract obstruction.

Despite only listing the above adverse reactions/risks, it is clear from the testimony of senior Ethicon Employees in both the Medical Affairs and Regulatory Affairs departments that every adverse reaction/risk that Ethicon is aware of today, it knew about at the time the TVT was first sold, marketed and launched.¹⁶⁸ Medical Director, Dr. Piet Hinoul, testified that Ethicon was aware that the following adverse events are associated with the TVT from the time the TVT was first sold¹⁶⁹:

- Erosions through the vaginal epithelium
- Infection
- Pain
- Urinary Problems
- Erosions that can decrease a patient's quality of life
- Dyspareunia including chronic dyspareunia
- Need for additional surgeries
- Need for removal of the device
- Urinary Tract Infections
- Dysuria
- DeNovo Urgency
- Mesh Exposure
- Fistula Formation
- Hematoma
- Abscess Formation
- Narrowing of the vaginal wall

¹⁶⁸ Beath, 7/12/13,608:13-20; Hinoul, 6/27/13, 551:12-552:9; Arnaud, 7/19/13,114:21-127:1

¹⁶⁹ Hinoul 7/27/ 13 542:11-582:13.

Erosion which can occur at any time in the future
 Contracture of the mesh causing pain
 Complications making it impossible to have sexual relations
 Worsening Incontinence

Yet, virtually none of these were in the TVT IFU at launch. Additionally, Catherine Beath, VP of Quality of Assurances and Regulatory Affairs, when discussing the October 20, 2008 FDA Public Health Notification (PHN),¹⁷⁰ testified that Ethicon was aware of all of the risks outlined in the PHN at the time of the launch of the TVT line of products. In other words, Ethicon had knowledge of all of the risks listed in the 2008 PHN at the time it launched the TVT.¹⁷¹ Medical Director, Dr. Weisberg, testified that Ethicon did not include: “permanent, lifelong, worsening and debilitating pain”, lifelong risk of surgical repairs for erosions, “severe or chronic inflammation“, collapse under strain and cause fibrotic bridging, that the product can degrade, that polypropylene is cytotoxic, severe erosion, or particle loss in the IFU.¹⁷² Former Medical Director, Dr. David Robinson, testified that Ethicon never informed physicians that patients may require multiple surgeries to treat erosions, that erosions could be severe and untreatable, and that patients could endure lifelong severe pain or dyspareunia. This is true despite, as discussed above; Ethicon had scientific knowledge of the risks at the time of launch.

Interestingly, in 2008, 2011 and 2012, Ethicon added numerous adverse reactions and risks to its Patient Brochures that were never disclosed in previous versions of the Patient Brochures. For some reason, though, these adverse reactions and risks have never been disclosed in the TVT IFUs. These risks are as follows:

¹⁷⁰ Eth.Mesh 07937826 2008 FDA Health Notification.

¹⁷¹ Beath 7/11/13 233:25-234:8, 245:21-246:1.

¹⁷² Weisberg Dep. (8/9/13) 968:12-972:21.

From Patient Brochures (never in IFU)

2008

Difficulty urinating

Pain

Scarring

Mesh Exposure requiring treatment

2011

Mesh exposure into the vaginal canal

Mesh exposure associated with pain during intercourse for the patient and partner¹⁷³

Mesh exposure which may require removal of exposed mesh in the office or operating room

2012

Pelvic Pain

Development of Urinary Incontinence

Voiding Difficulties

Hemorrhage or hematoma

Urinary tract infection

Wound healing problems

Injury to ureters

Pelvic abscess formation

Risk of infection

Vaginal scarring

Mesh contracture (mesh shortening due to scar tissue)

For a surgeon to properly inform the patient of all the known risks involved in any procedure involving an implantable medical device, the surgeon relies upon the manufacturer to be aware of and convey all characteristics of its products that could impact safety and efficacy. Specifically, surgeons rely on the “Adverse Events/Risks” section of a medical device IFU to gain knowledge regarding adverse events or undesirable effects that the company knows are associated with the product.

If you compare the adverse reactions/risks in the TVT IFUs to the adverse reactions/risks Ethicon knew at the time of the launch of TVT, it is clear that there are numerous adverse

¹⁷³ It should be noted that this adverse reaction/risk was taken out of the 2012 Patient Brochure.

events/risks absent from the IFU that should have been included. For some reason, Ethicon chose to exclude from the IFU multiple adverse reactions and risks that even Ethicon's own employees in senior management admit should have been included.

Even though Ethicon changed its Patient Brochures in 2011 and 2012 to include additional significant adverse events/risks, it never added the same information to the TVT IFU. This is true despite the fact that Ethicon had internal discussions about updating the IFU in 2009 after the 2008 FDA Public Health Notification (PHN). Specifically, a meeting was held and one of the purposes of the meeting was to decide if the risks of tape exposure and post-operative dyspareunia following TVT implant should be included in the IFU.¹⁷⁴ After discussing the 2008 PHN, competitors labels and Remetrex issues, the impressions of those at the meeting were that tape exposure/erosion/extrusion was very frequently reported, that patients did not feel there was adequate pre-op consent or risk-benefit assessment before the TVT was implanted, that patient specific concerns about exposure/erosion/extrusion, incontinence recurrence, post-operative dyspareunia and pain were affecting the quality of life and daily routine of patients implanted with the TVT, and that re-operations and post-operative complications were disproportionate to women's pre-operative-consent-expectations.¹⁷⁵ This was a meeting attended by both regulatory and medical affairs.

Ethicon's Associate Medical Director of Worldwide Customer Quality Meng Chen, M.D., Ph.D., was responsible for asking for this meeting and encouraged her superiors to consider updating the Adverse Reactions section of the IFU. She testified that she reviewed a "couple of thousand complaints" as Associate Medical Director and started to have concerns

¹⁷⁴ Eth.Mesh 04081189

¹⁷⁵ Eth.Mesh 04081189

about complaints related to mesh exposure and dyspareunia.¹⁷⁶ Dr. Chen also testified that she was concerned that patients did not seem to understand the risks associated with the TVT before the implant.¹⁷⁷ Finally, she was concerned and told her superiors that patients felt that they were not being properly consented about the risk of erosions, recurrent erosions and the risk of dyspareunia.¹⁷⁸ It was for these reasons that Dr. Chen felt that the IFU was inadequate and needed to be updated.

Dr. Chen testified that she spoke with numerous patients about their specific concerns that they “did not feel there were adequate pre-op consent or risk-benefit assessment” before undergoing TVT surgery.¹⁷⁹ She added that patients told her that if they had been fully informed of the risks associated with the TVT, they would have made a different decision or would have chosen “something different.”¹⁸⁰

On December 19, 2008, after Dr. Meng Chen had received a complaint from a number of patients about not being fully informed of the risks of the procedure, she recommended to senior management that the IFU be updated:

[The patient] was given the most accurate consent for the potential adverse reaction known in 2005. However, we are in 2008 now, and there are two more TVT family products (TVTO and TVTS) on the market. Our post-market knowledge with these products are much more than what we have in the IFUs of all three types of TVTs (TVT-Abdominal, Obturator and Secur). My reason for bringing this point to you is maybe you may look into it from senior management perspective and to facilitate the IFU update for all three TVTs, particularly in the area of ‘Potential Adverse Reactions’ One of the paths for a better pre-operative consent is to provide an updated IFU to the operating physicians that reflecting the current knowledge of the manufacturer’s on the potential adverse reactions.¹⁸¹

¹⁷⁶ Chen, 10/30/13, 240:6-13

¹⁷⁷ Chen, 10/30/13, 242:9-14

¹⁷⁸ Chen, 10/30/13, 243:16-244:17

¹⁷⁹ Chen, 10/30/13, 231:20-232:5

¹⁸⁰ Chen, 10/29/13, 176:20-177:4

¹⁸¹ Eth.Mesh.04092868.

In addition, in a January 29, 2009 email, Dr. Chen wrote again that the IFU should be updated to make it clear that the irritation and foreign body response were a result of the tape itself and that this “could result in tape extrusion, tape erosion, fistula formation or inflammation.”¹⁸² Dr. Chen testified that if she knew of these risks (post-op dyspareunia, mesh exposures, erosions) in 1999 around the time the product was launched, she would have recommended that the IFU been updated so patients knew about these risks.¹⁸³ In her opinion, a responsible medical device company with a responsible medical director would have updated the IFU.¹⁸⁴

Dr. Chen wasn’t the only employee at Ethicon that believed the TVT IFU was inadequate and needed to be updated. When working on the Mini-O/Abbrevio IFU, Ethicon employees noted that the older IFU’s should be updated. Dr. Aaron Kirkemo wrote:

I would agree from the meeting today that now that we have 12+ years of experience with TVT classic that learnings from the field would probably drive a relook at the TVT Classic IFU as reflected by some of your comments in this document.¹⁸⁵

In response, Dr. Robinson asked: “has there been agreement re: a project to revise TVT and TVTO?”¹⁸⁶ There was indeed agreement at upper management – there would be no revision to incorporate what they had learned: “Per Scott C and Stale, they just want to “look forward” with this project. Their plans are to leave TVT Classic [and TVT-O] as is. Aaron.”¹⁸⁷

Despite these discussions and Ethicon’s knowledge of these serious, devastating and life-changing adverse events/risks, to this day, Ethicon has never updated or changed its IFU to include important safety information. The bottom line is that Ethicon has clearly failed to

¹⁸² Eth.Mesh.04094863 (e-mail from Dr. Meng Chen to Bryan Lisa, Jan. 29, 2009).

¹⁸³ Chen, 10/30/13, 244:5-246:13

¹⁸⁴ *Id.*

¹⁸⁵ ETH.MESH.01239065 at 9066 (July 14, 2009 email from Aaron Kirkemo MD to Piet Hinoul MD and David Robinson MD).

¹⁸⁶ *Id.*

¹⁸⁷ *Id.*

adequately warn physicians about significant adverse events and risks associated with the TVT, including permanent, lifelong and debilitating pelvic pain, lifelong sexual complications and dysfunction (including but not limited to dyspareunia), worsening incontinence, the lifelong risk of multiple surgeries, the need for removal of the device, the lifelong risk of erosions, and the risk of serious complications that could negatively impact a patient's quality of life. For this reason, Ethicon's IFU is, and has always been, inadequate and defective. Furthermore, Ethicon deviated from the standard of care required of a reasonable medical device manufacture by failing to adequately disclose these known adverse reactions and risks to physicians and, as a result, has denied physicians and patients the ability to make an informed choice regarding the use of the TVT.

2. Several of the adverse reactions, risks, and characteristics of the mesh Ethicon included in the IFU are inaccurate or inappropriately downplayed

Not only has Ethicon failed to include many adverse reactions and risks in its IFU, but Ethicon has also inappropriately downplayed several adverse reactions and risks that are included in the IFU. This is especially true with respect to erosions. The Adverse Reactions section of the TVT IFU (which has been in place from September, 2000 until present day), states as follows:

* Transitory local irritation at the wound site and a transitory foreign body response may occur. This response could result in extrusion, erosion, fistula formation and inflammation. (emphasis added)

This language significantly downplays the permanent nature of erosions and suggests to physicians that erosions are a "transitory" or a temporary problem. The inadequacy of this statement about erosions was the subject of discussion among the medical directors in Ethicon many years ago. In fact, Dr. Chen, the same company doctor who was charged with assessing complaints from women and doctors related to the TVT and who told her superiors that the

warning label is inadequate, also told her superiors that the “transitory” statement about erosions is misleading. Specifically, in an email exchange between Dr. Chen, and Bryan Lisa, an employee in the Regulatory Affairs Department on January 29, 2009,¹⁸⁸ Dr. Chen makes it clear to Bryan Lisa that calling erosions “transitory” is not consistent with her assessments of the complaints when she says:, “Pardon me again, from what I see each day, these patient experiences are not “transitory” at all.”¹⁸⁹

Ethicon employees admit that erosions are not transitory. Ethicon also knew that erosions could occur many years after implantation of the device. In Minutes from Ethicon’s June 22, 2001 Scientific Advisory Committee on Pelvic Floor Repair, it was the consensus of the group that “Erosion is a risk. Erosion, possibly an infection response. Typically seen by 3 mos, usually by 6-12 mos. Can present late, 3 years. To vagina-not a good situation. To bladder, urethra or rectum-a very bad situation.”¹⁹⁰ Nevertheless, Ethicon decided to downplay the risk of erosions in the IFU --a decision that is entirely consistent with the recommendation Dr. Axel Arnaud, European Science Director at Ethicon, gave to Dr. Weisberg, the Medical Director in the United States when he told him to “be more elusive” when describing erosions as a complication in a foreign regulatory document.¹⁹¹ In fact, based on my review of the testimony and evidence in this case, Ethicon consistently (and inappropriately) downplayed the severity and rate of erosions in communications with physicians.¹⁹²

There are other examples of Ethicon downplaying the risks of adverse reactions in the TVT IFU. For example, the IFU states that the mesh can “potentiate infection” when, in fact, Ethicon has known at all times while selling the TVT that the TVT can actually cause a new

¹⁸⁸ Eth.Mesh.04093125 1/29/2009 Email between Meng Chen and Bryan Lisa.

¹⁸⁹ Id.

¹⁹⁰ Eth.Mesh.02089392.

¹⁹¹ Eth.Mesh.03910175-03910177.

¹⁹² ETH.MESH.00339438, ETH.MESH.00658059

infection that did not already exist. If by “potentiate”, Ethicon means “exacerbate an existing infection”, then the statement is misleading at best. Ethicon had a duty to warn physicians in its IFU that a slimy, protective biofilm could form on the mesh leading to painful erosions, recurrent, late infections and the need for mesh removal. By not doing so, they did not adequately warn physicians about these important risks or, by extension, provide surgeons with an opportunity to discuss these risks with their patients.

As stated above, Ethicon also inappropriately downplayed the risk of the chronic foreign body reaction associated with the TVT mesh in the TVT IFU. The IFU for the TVT states: “a transitory foreign body response may occur.”¹⁹³ This statement is simply untrue. Multiple senior medical directors at Ethicon admit that this statement is inaccurate.¹⁹⁴

Finally, as discussed more fully above, Ethicon choose to state in its IFU for the TVT that the mesh does not degrade (“is not absorbed, nor is it subject to degradation or weakening by the action of enzymes”), when it cannot be disputed that the mesh does, in fact, degrade.¹⁹⁵ Ethicon’s internal degradation experts agree.¹⁹⁶ Ethicon made this misrepresentation in the IFU despite the scientific literature, its own internal studies, and reports from consultants it hired that state that degradation of the mesh does, in fact, occur.¹⁹⁷

It is my opinion that the above mentioned inaccuracies and misleading statements in the TVT IFU misled physicians about the safety of the TVT and, in turn, led to the adoption of the TVT device when other safer alternatives existed. Because of the inadequate warnings and the inaccuracies in the TVT IFU, patients were also deprived of important safety information that

¹⁹³ ETH.MESH.05225354

¹⁹⁴ Hinoul, 1/14/2014, 1192:4-1199:12; Robinson 9/11/13, 1087:7-1089:15

¹⁹⁵ Eth.Mesh.05225354; Eth.Mesh.02340306; Eth.Mesh.02340471; Eth.Mesh.05222673; Eth.Mesh.02340504; Eth.Mesh.03427878.

¹⁹⁶ Deposition of Thomas Barbolt, M.D., January 8 2014, pg 409; 516-17

¹⁹⁷ See Footnotes 20-30 supra.

they needed in order to make an informed decision about whether to have the TVT mesh permanently implanted in their bodies.

3. Ethicon failed to include warnings about patients at increased risk upon implantation of the TVT and a warning that use of general anesthetic increases the risk to women

Finally, Ethicon also failed to include pertinent warnings in its IFU about the increased risks to women with certain pre-existing conditions or a warning that if the TVT procedure was performed under general anesthetic rather than local anesthetic (as recommended by the Professor Ulmsten, the inventor of the TVT), the efficacy of the TVT decreases and the risk to women increases. Once again, this is true, even though Ethicon knew that obese women or women with certain pre-existing conditions were being implanted with the TVT and that general anesthetic was commonly used during the insertion of the TVT.

Specifically, the TVT IFU needed to include the following warnings:

WARNINGS

Obese patients, elderly patients, and younger, active women may have less successful outcomes following TVT implantation.

The risk of mesh extrusion is increased in women with postoperative infection, previous vaginal surgery, vaginal atrophy or vaginal injury.

Performance of this procedure with general anesthesia increases the risk of urinary retention, erosion and failure of the surgery.

Ethicon promoted the TVT as a reproducible technique that was appropriate for all patients, including obese and elderly patients.¹⁹⁸ For example, Ethicon instructed its sales force to specifically target physicians to use the TVT in obese patients.¹⁹⁹ However, as Ethicon's Medical Director, Dr. Kirkemo, testified, obese patients do not fare well with these devices:

¹⁹⁸ ETH.MESH.00339438, ETH.MESH.00658058, ETH.MESH.10216874

¹⁹⁹ See, e.g., ETH.MESH.00640394 (trying to convince physicians to use TVT-O on obese patients); ETH.MESH.05119622 at 9623 (TVT "is a good choice for the obese patient or elderly patient....").

Q. One of the things that was actually shown in the TVT World study that you worked on was that for obese patients, for example, the efficacy was significantly down when slings were used in obese patients; is that correct?

A. Obese people tend -- not to do as well.

In fact, Ethicon's study showed obese patients had about one half the success of those patients who were not obese. In addition, as Dr. Kirkemo testified, obese women suffered from more complications: "Their chance of success goes down. Their risk of complications goes up."²⁰⁰

Yet, Ethicon did not put this critical information into the IFU. Dr. Kirkemo testified further:

Q. Did you ever put that in the IFU?

A. No....²⁰¹

Not only did Ethicon not put this critical information in the IFU, Ethicon also did not inform patients:

Q. Did you ever tell patients that in a single patient brochure, that if they were obese, their chances of this being successful were less than half?

A. We did not.²⁰²

Ethicon also did not include information in its IFU about how the TVT had less efficacy and higher risk for older women or younger, active women.

Q. Did you -- you also learned in the TVT World study, or maybe you knew this before, too, that being elderly decreased, or being very young, in fact, decreased the efficacy of the Ethicon sling procedures; correct?

A. With any incontinence operation, old people tend not to, you know, do as well.

Q. And was that ever put in a patient brochure or communicated to patients as far as you know?

A. As near as I can tell, in any marketing document, no.

Q. And what about the very young or the younger women; that was shown in TVT World that even younger women had lower efficacy; correct?

A. Some women that are very, very active can -- and have ISD can overcome the effect of the sling.

Q. In other words, the sling can fail.

A. The sling can be less than a hundred percent effective.

Q. And that was never actually communicated to patients as far as you know, correct, by Ethicon?

A. To my knowledge, no.

²⁰⁰ Kirkemo Dep. (1/7/2014) 556:24-557:1.

²⁰¹ Kirkemo Dep. (1/7/2014) 556:4-19.

²⁰² Kirkemo Dep. (1/7/2014) 557:5-557:9.

Q. And neither the older women or the younger women in issue we were just talking about, neither of those are included in the IFU; correct?

A: Those specific things are not mentioned.²⁰³

Ethicon knew that there were other patient populations that also faced increased risk or lower success rates with the TVT. Specifically, Ethicon knew that women who had prior pelvic surgery, prior pelvic injury or an infection, could be at increased risk if undergoing the TVT surgery. In fact, Ethicon discussed as early as 1999 putting another warning in the TVT IFU related to patients who had previous surgeries because of scar tissue.²⁰⁴ Ethicon was concerned that the risk of mesh extrusion was increased in women with postoperative infection, previous vaginal surgery, vaginal atrophy or vaginal injury.²⁰⁵ Dr. Isenberg, Ethicon medical director, admitted that if Ethicon knew this, it would have been reasonable to include a warning and, further, physicians and their patients would want to know this information. Again, despite these discussions in 1999 and former Medical Director Dr. Isenberg's opinions that it would be reasonable to have this information in the IFU, to this day, this critical information remains absent from the IFU.

Finally, Ethicon also knew that the method of anesthesia utilized during the TVT surgery could affect patients' outcomes, but didn't disclose that information to physicians or patients. Ethicon's internal documents, including interviews with Ethicon's key opinion leader, Dr. Carl Nilsson, Ethicon U.S. Marketing Research documents, and letters from the inventor of the TVT (Dr. Ulmsten) show that Ethicon knew that performing the TVT procedure under general anesthesia as opposed to local anesthesia decreased the chance for success of the surgery and

²⁰³ Kirkemo Dep. (1/7/2014) 557:10-558:21.

²⁰⁴ ETH.MESH.08505071, ETH.MESH.00203456, Eth.Mesh.00159634-00159719 at 00159697

²⁰⁵ Id.

also increased a patient's risk of urinary retention and erosions.²⁰⁶ This is further supported by the testimony of Dr. Richard Isenberg, a former medical director for Ethicon, who was at Ethicon just after the initial launch of the TVT.²⁰⁷ Dr. Isenberg testified that the IFU could be better worded so that physicians knew that local anesthesia should be preferred over general anesthesia.²⁰⁸ In addition, according to Dr. Isenberg, Dr. Ulmsten, inventor of the product, informed Ethicon that the TVT procedure should be carried out under local anesthesia unless it was a special situation.²⁰⁹ Despite the inventor's desire to have this language listed, to this day, it does not appear in the IFU.²¹⁰ Dr. Isenberg was also aware that using general anesthesia could cause the success rate of the procedure to go down and put the patient at increased risk for urinary retention and erosions.²¹¹ He testified that he believes a responsible company should have put this information in the IFU because the IFU is the one document that you can count on every physician receiving.²¹² I agree. Again, however, to this day, this warning does not appear in the TVT IFU.

The TVT is dangerous and can cause significant, lifelong injury to women, due in part to its "one-size fits all" design. Ethicon failed to inform physicians that there are certain patient populations that face greater risks and less success with the TVT. Ethicon needed to pass this critical information on to physicians in the IFU so that they could have an appropriate informed consent discussion with their patients.

In summary, Ethicon not only failed to disclose adverse reactions, risks, and failed to warn about known groups of women who would be at increased risk due to pre-existing

²⁰⁶ Eth.Mesh.04048515-04048520; Eth.Mesh.00130934-00130941, Eth.Mesh. 00400954-00400956

²⁰⁷ Isenberg, 11/6/13, 461:16-530:13

²⁰⁸ Id. at 526:25-528-18

²⁰⁹ Id. at 553:15-554:21

²¹⁰ Id.

²¹¹ Id. at 566:9-15

²¹² Id. at 566:3-8

conditions or the way the procedure was performed, but also downplayed significant safety information and risks in the IFU. This is information physicians need to know in order to have a fair and proper conversation with their patients about the use of the TVT. Physicians rely on device manufacturers to inform them of the risks and complications associated with its products instead of downplaying them or omitting them altogether. By not disclosing this safety information to physicians and their patients, Ethicon failed to act like a reasonable and prudent device manufacturer.

D. Ethicon did not disclose to physicians in its IFU information regarding characteristics of Ethicon's old construction TVT mesh (Prolene) that make it potentially dangerous and unsuitable for its intended application as a permanent prosthetic implant for stress urinary incontinence, including that it is small pore, heavy weight mesh, it degrades over time, causes chronic foreign body reactions, fibrotic bridging, mesh contracture/shrinkage, and fraying, sharp edges, particle loss, roping and curling, and that it deforms with tension causing the pores to collapse.

As discussed above, the goal of the IFU is to communicate the most important safety information and risks attributable to the TVT device. This includes information related to characteristics of the product that can lead to injury or harms to the patients who will use the product. In this case, it cannot be disputed that all known hazards associated with the TVT should have been included in the IFU at all times. In fact, Ethicon's own medical and regulatory experts agree that an IFU should never exclude known hazards or complications.²¹³ Unfortunately for physicians and patients, though, Ethicon has never disclosed multiple hazardous conditions/characteristics associated with the Prolene mesh in the TVT which it knows can lead to very serious injuries in women.

As previously discussed, Ethicon knew about characteristics of the old construction Prolene mesh that could lead to significant harms and hazards for patients. . Ethicon knew that

²¹³ Weisberg, 8/9/13, 959:19-960:15.

women could suffer harms and hazards due to the fact that the Prolene mesh is small pore, heavy weight mesh, it degrades over time, causes chronic foreign body reactions, fibrotic bridging, contracture/shrinkage, and because it frays, has sharp edges, loses particles in women's vaginal tissues, ropes, curls, deforms and the pores collapse with very little tension.²¹⁴ However, despite its knowledge regarding these hazardous characteristics, Ethicon never disclosed this information to physicians in its IFU. As a result, physicians have not been able to have a fully informed conversation with their patients about the safety of the TVT device.

In fact, there is clearly a misconception among physicians who use midurethral slings, including the TVT, that the mesh used is the safest alternative mesh when, in fact, Ethicon internal documents make it clear that is not true. For example, Ethicon inappropriately markets the TVT mesh as "light weight, macroporous" mesh even though their internal documents say otherwise.²¹⁵ Because of Ethicon's marketing in this regard, there is a misconception among physicians that the TVT mesh is, in fact, light weight and macroporous. I also believe, based on my experience, my training, and all my work in this case, that very few, if any, physicians actually know that the TVT mesh degrades, loses particles, or ropes, curls and deforms with very little tension. Even if they do, though, they do not have the knowledge, like Ethicon does, that such dangerous characteristics can cause harm to their patients.²¹⁶ For this reason, it is imperative that Ethicon disclose these potentially dangerous characteristics and the complications/risks associated with each to doctors and patients. Ethicon has failed to do this

²¹⁴ I discussed each of these defects previously in some detail and, thus, will not rehash the same here.

²¹⁵ Cobb W, Kercher K, Heniford T. The Argument for Lightweight Polypropylene Mesh in Hernia Repair. *Surgical Innovation*. 2005; 12(1):T1-T7; ETH.MESH.08315779, ETH.MESH.05479535. ETH.MESH.04941016, ETH.MESH.04941016, ETH.MESH.05479411

²¹⁶ Moalli et al, Tensile properties of five commonly used mid-urethral slings relative to the TVT, *International Urogynecology Journal*, 2007 (stating that textile properties, including pore size and weight are provided by manufacturers).

and, as a result, physicians have not been able to make an informed decision about using the TVT and patients have been deprived of the information as well.

E. Ethicon failed to adequately describe, inform or explain to physicians how to properly “tension” the TVT and inform them that improper tension on the mesh decreased effective pore size and interfered with incorporation into tissue.

TVT stands for and has consistently been marketed by Ethicon as “Tension-free Vaginal Tape.”²¹⁷ Presumably, this means the mesh should be inserted under the urethra without tension. However, the term “tension-free” is misleading. In practice, too little or no tension results in failure to treat the underlying condition of urinary incontinence. On the other hand, as suggested by Ethicon’s own internal documents, too much tension can result in serious complications such as urethral erosion.²¹⁸

The IFU provides little guidance on proper tensioning of the TVT. Specifically, once the tape is placed, surgeons are simply instructed to pull the needles upwards “to bring the tape (sling) loosely, i.e. without tension, under the midurethra” and to “adjust the tape so that leakage is limited to no more than one or two drops.”²¹⁹ The IFU’s Warnings and Precautions section cautions surgeons to “[e]nsure that the tape is placed with minimal tension under the mid-urethra.”²²⁰ Yet in the very same section, the surgeon is instructed “to place the tape tension-free in the mid-urethral position” to minimize the risk of de novo detrusor instability.²²¹ Finally, the IFU’s “Adverse Reactions” section provides that “over correcting, i.e. too much tension applied to the tape, may cause temporary or permanent lower urinary tract obstruction.”²²² The IFU’s conflicting instructions with regard to tensioning of the tape, i.e. “without tension,” “with

²¹⁷ Smith 6/5/13, 1186:9-16.

²¹⁸ Eth.Mesh.05529274; Eth.Mesh.04044797; Eth.Mesh.05529653; Eth.Mesh.00161131.

²¹⁹ Eth.Mesh.05222686, emphasis added.

²²⁰ Eth.Mesh.05222687, emphasis added.

²²¹ Eth.Mesh.05222567, emphasis added.

²²² Eth.Mesh.05222687, emphasis added.

minimal tension,” “tension-free” and “overcorrecting, i.e. too much tension” are clearly confusing and inadequate despite the fact that Ethicon knew as early as 2000 that improper tensioning could lead to complications and, therefore, the IFU needed to be “clear.”²²³

Ethicon recognized as far back as November 1999 that TVT tension adjustment was considered "high need" and surgeons had a hard time sticking to proposed technique.²²⁴ By 2000, Ethicon recognized that excess tensioning during initial placement could create a risk of erosion.²²⁵ In an email dated February 13, 2001, Medical Director Axel Arnaud wrote “there is clearly a need for standardization of the TVT procedure to avoid excessive tension on the mesh. We should aggressively work in order to develop a product and I would like to take the responsibility for this.”²²⁶ In May 2002, Axel Arnaud continued to recognize the need to develop a safer device “in order to prevent excess tension of the tape.”²²⁷ In 2003, Ethicon recognized that a challenge with the TVT procedure remained complications “associated with over-tensioning of the sling and the inability to obtain precise biofeedback and adjustment during and/or after the procedure.”²²⁸ Indeed, Dr. Nilsson, the “father of the TVT”, discussed that the TVT done under general anesthesia with a cough test was 70% successful compared to a 85% success rate when done with local anesthesia and a cough test.²²⁹

The lack of clear direction on tensioning in the IFU is demonstrated in September 2004 emails from Sales Representative Shannon Campbell in which she writes: “What is a huge challenge to a rep trying to make this right, is that we really don't know what the right amount [of tensioning] is. We know this is a quick fix to the problem, but not a clinically backed

²²³ Eth.Mesh.01317523 .

²²⁴ Eth.Mesh.05641096 .

²²⁵ Eth.Mesh.05529274; Eth.Mesh.04044797; Eth.Mesh.05529653; Eth.Mesh.00161131.

²²⁶ Eth.Mesh.03915380.

²²⁷ Eth.Mesh.03907468.

²²⁸ Eth.Mesh.00259271.

²²⁹ Eth.Mesh.04048515 at Eth.Mesh.0408516 7/01/08 KOL Interview: Carl G. Nilsson, Project Scion

solution. It's almost like trying to decide if a 8, 10, or 12 mm Hagar dialator is best for tensioning TVT with the patient under general. We learned the cough test, but relied on surgeons experience with the tensioning under general.... This has been such a gray area and everyone seems to have their own tensioning technique.” She continues: “I feel I got a little grilled over my suggestion of tensioning, yet there is no clear direction on tensioning. I’m not a rebel looking for my own way of doing this. I’m a rep trying to figure out what is best from my experience with surgeons and what I see the product doing in the OR. ...The reason for my question is to see if someone had the proper wording we need to use as rep’s that eliminates our liability with the product in the OR concerning tensioning.”²³⁰

In December 2006, Ethicon Marketing Director Allison London-Brown referred to tensioning as a “sticky” question and acknowledged that “we cannot accurately describe [tensioning] in writing.”²³¹ Meanwhile, Ethicon knew that patients were suffering from erosions and, in fact, would often blame the physician as the cause of the erosion for putting “too much tension on the device.”²³² At least by 2007, it seems Ethicon finally acknowledged that “TVT has never been tension free!” despite years of marketing it otherwise.²³³ For example, in 1999, Ethicon utilized marketing pieces for “TVT Tension Free Vaginal Tape” which claimed “Tension-free Support Only When Needed” which “reduces possibility of urethral erosion.”²³⁴ A 2001 marketing piece for “Gynecare TVT Tension-Free Support for Incontinence” claimed “most complications are minor and are avoidable with adherence to procedural technique and instructions for use.”²³⁵ In 2004, during the same time period when Shannon Campbell was lamenting the problems with tensioning, Ethicon continued to promote TVT as “the leader in

²³⁰ Eth.Mesh.00864503.

²³¹ Eth.Mesh. 01784428-01784435

²³² Eth.Mesh.02625055, Eth.Mesh.02627811, Eth.Mesh.02625375, Eth.Mesh.02625155

²³³ Eth.Mesh.06861473.

²³⁴ Eth.Mesh.00161444.

²³⁵ Eth.Mesh.00339437.

midurethral sling devices” for “tension-free support for incontinence.”²³⁶ Even after Ethicon acknowledged that TVT has never been tension free, the company continued to market it as “Tension-free Support for Incontinence.”²³⁷

Physicians were also not informed in Ethicon’s product IFU that tension on the mesh arms decreases effective pore size and interferes with incorporation into tissue. Engineer Christophe Vailhe testified that “excessive uniaxial tension on the mesh will decrease the pore size and lead to poor tissue integration.”²³⁸ In addition, Mr. Vialhe testified that “excessive tension on the mesh would lead to the decrease in pore size that can lead to poor tissue integration”²³⁹ Engineer Dan Burkley also testified that once the TVT Prolene mesh is either stretched by the surgeon or stretched by in-vivo due to forces in a women’s body, it can alter the structure of the pores.²⁴⁰

The IFU failed to adequately instruct surgeons on the critical subject of tensioning as repeatedly acknowledged by Ethicon. Ethicon now claims that “tension-free” does not really mean tension-free, but rather, means less tension than as seen in the Burch procedure.²⁴¹ Yet, despite its awareness of the problems associated with tensioning, Ethicon failed to revise the conflicting and ambiguous IFU to provide adequate direction on the proper amount of tensioning even though Ethicon was fully aware that improper tensioning could lead to serious complications such as urinary retention, voiding difficulties, de-novo detrusor instability, dyspareunia, vaginal extrusion and urethral erosion.

Ethicon failed to act as a reasonable and prudent medical device manufacturer by failing to inform physicians how to properly tension TVT and that improper tension could affect the

²³⁶ Eth.Mesh.00160813 .

²³⁷ Eth.Mesh.00164643; Eth.Mesh.00339053 .

²³⁸ Vailhe, 6/20/13, 224:10-226:21

²³⁹ Vailhe, 6/20/13, 224-226

²⁴⁰ Burkley 5/22/13 430:3-431:10.

²⁴¹ Smith 6/4/13 524:20-525:13.

pore size of the mesh. These failures by Ethicon have resulted in numerous injuries to patients, including, but not limited to urinary retention, voiding difficulties, de-novo detrusor instability, dyspareunia, and vaginal extrusion and urethral erosion.

F. Ethicon did not inform physicians and their patients that Material Safety Data Sheets (MSDSs) for polypropylene resin used to manufacturer polypropylene meshes warned against use of the mesh in a permanently implanted medical device and that studies show that polypropylene causes sarcomas in laboratory rats.

According to Ethicon Medical Director, Dr. Martin Weisberg, a Material Safety Data Sheet (MSDS) is “a document that discusses the product, the composition, any potential hazards from it . . . Generally, the safety particular of products.”²⁴² As it relates to polypropylene, I have reviewed several MSDSs for polypropylene resin used to manufacturer meshes used in various pelvic floor meshes. All of the MSDSs discussed below are available to the public.

Sunoco, the manufacturer for the polypropylene resin used to manufacture Ethicon’s pelvic floor products lists the possibility that polypropylene mesh can cause tumors or cancer. This is documented by the Sunoco MSDS²⁴³ from April 13, 2005 which states in relevant part:

15. OTHER INFORMATION

Follow all MSDS/label precautions even after container is emptied because it may retain product residue.

COMPONENT TOXICITY: Polypropylene has been tested in laboratory rats by subcutaneous implantation of discs or powder. Local sarcomas were induced at the implantation site. No epidemiological studies or case report suggest any chronic health hazard from long term exposure of polypropylene decomposition products below the irritation level. (OARC, 19, 128).²⁴⁴

²⁴² Weisberg 8/9/13 909:2-9.

²⁴³ Eth.Mesh.02026591-02026595.

²⁴⁴ Id. at 02026595.

Dr. Martin Weisberg, Ethicon Medical Director, is not only familiar with this MSDS, he also has personal experience with it. Dr. Weisberg agrees that the manufacturer of Ethicon's mesh did a study by implanting it under the skin of rats and it did in fact induce sarcomas.²⁴⁵ Dr. Weisberg also agrees "if there was evidence of cancer-causing abilities of polypropylene . . . a reasonable doctor would want to know."²⁴⁶ And, despite evidence to the contrary in the above MSDS for the resin used to make the polypropylene mesh for TVT, he is not aware of any instance when Ethicon "disclosed to any doctor that there's any evidence that the use of polypropylene mesh might induce sarcomas in its patients."²⁴⁷

Dr. David Robinson, a former Ethicon Medical Director, testified that Ethicon never performed any studies or research to determine whether polypropylene could cause cancer in the long term.²⁴⁸ In addition, he testified that Ethicon never disclosed "the potential that polypropylene in the product could be cancer causing."²⁴⁹ Dr. Robinson also testified that it would be reasonable for physicians to want to know about polypropylene possibly causing cancer.²⁵⁰

Another MSDS from Chevron Phillips,²⁵¹ a manufacturer of polypropylene resin states:

MEDICAL APPLICATION CAUTION: Do not use this Chevron Phillips Chemical Company LP material in medical applications involving permanent implantation in the human body or permanent contact with internal body fluids or tissues.

Do not use this Chevron Phillips Chemical Company LP material in medical applications involving brief or temporary implantation in the human body or contact with internal body fluids or tissues unless the material has been provided directly from Chevron Phillips Chemical Company LP under an agreement which expressly acknowledges the contemplated use.

²⁴⁵ Id. at 930:3-8

²⁴⁶ Id. at 951:6-10

²⁴⁷ Id. at 951:11-16

²⁴⁸ Robinson 9/11/13, 1107:12-18

²⁴⁹ Robinson 9/11/13, 1114:15-18

²⁵⁰ Robinson 9/11/13, 1115:5-12

²⁵¹ Chevron Materials Safety Data Sheet Marlex Polypropylenes (All Grades) Revision Number: 3 (T-3137).

Chevron Phillips Chemical Company LP makes no representation, promise, express warranty or implied warranty concerning the suitability of this material for use in implantation in the human body or in contact with the internal body fluids or tissues.

With respect to the Chevron Phillips MSDS, Ethicon Medical Director, Dr. Martin Weisberg, testified that he did not have the Chevron Phillips MSDS in 2001 when he reviewed the Sunoco MSDS and no one at Ethicon alerted him to it.²⁵² If he had been alerted to the Chevron Phillips MSDS, it may have “triggered” an investigation on his part.²⁵³ He also believes that if Ethicon knew about this MSDS, Ethicon should have studied the issue and, if they did not do so, it would have been a violation of the company Credo.²⁵⁴

Total Petrochemicals, the polypropylene resin manufacturer for the polypropylene used in AMS’ pelvic floor products, Technical Data Sheet for Polypropylene PPR 7220, states in bold red lettering

“Under no circumstances are any products sold by Total Petrochemicals suitable for human or animal implants.” It is further documented that, “The above-mentioned product is NOT in compliance with the US pharmacopoeia because we DID NOT perform required tests.” (emphasis from the original document).

The manufacturer of the polypropylene resin for the polypropylene used in competitor pelvic floor products, Phillips Sumika Polypropylene Company, included a similar warning in its MSDS²⁵⁵. Specifically, it states:

"Do not use this Phillips Sumika Polypropylene Company material in medical applications involving permanent implantation in the human body or permanent contact with internal body fluids or tissues. Do not use Phillips Sumika Polypropylene Company material in medical applications involving brief or temporary implantation in the human body or contact with internal body fluids or tissues unless the material has been provided directly from Phillips Sumika

²⁵² Weisberg, 8/9/13,944:16-945:5.

²⁵³ Id.

²⁵⁴ Id. at 947:4-19

²⁵⁵ Phillips Sumika Polypropylene Company Material Safety Data Sheet Marlex Polypropylene (All Grades) Revision Number: 5.03 Revision Date: 12/4/2008

Polypropylene Company under an agreement which expressly acknowledges the contemplated use. Phillips Sumika Polypropylene Company makes no representation, promise, express warranty or implied warranty concerning the suitability of this material for the use in implantation in the human body or contact with internal body fluids or tissues."

As discussed above, the possibility that polypropylene mesh can cause tumors or cancer is documented in the Sunoco MSDS, the manufacturer of the polypropylene resin used in the TVT Prolene mesh.²⁵⁶ Specifically, the Sunoco MSDS from April 13, 2005 states: COMPONENT TOXICITY: Polypropylene has been tested in laboratory rats by subcutaneous implantation of discs or powder. Local sarcomas were induced at the implantation site. No epidemiological studies or case report suggest any chronic health hazard from long term exposure of polypropylene decomposition products below the irritation level."²⁵⁷

Despite this warning in the MSDS for the polypropylene resin used to manufacture the TVT mesh, there is no evidence that Ethicon informed surgeon about this important information contained in various Manufacturer Safety Data Sheets (MSDS) regarding the use of polypropylene. This information includes the dangers of using polypropylene in a permanent implanted medical device set forth in MSDS that were in the public domain and available to Ethicon if they chose to look. Ethicon also failed to inform physicians that laboratory studies on rats showed that polypropylene caused sarcomas.

The fact that this information has not been disclosed to physicians in any manner (IFUs, direct letters or promotional materials) is especially concerning in light of literature showing reports of cancer associated with polypropylene. Specifically, there have been cases of pseudotumor reported in polypropylene for hernia mesh²⁵⁸ and inflammatory myofibroblastic

²⁵⁶ Eth.Mesh.02026591-6595.

²⁵⁷ Eth.Mesh.02026595.

²⁵⁸ Kassem, M., Community Oncology, Volume 7/Number 4/April 2010; Mehrotra, PK, Hernia, 2006 10:192-194

tumor of low malignant potential with a TVT device.²⁵⁹ In addition, there have been 2 cases of bowel cancer associated with mesh used for abdominal sacralcolpopexy, one associated with mersilene and one with polypropylene and TVT placement.²⁶⁰ A case of primary vaginal leiomyosarcoma associated with TVT and anterior repair with Bard Duraderm has also been reported.²⁶¹

Finally, a report of angiosarcoma associated with Darcon vascular grafts was reported in 1999.²⁶² The authors of this article noted at least 8 other sarcomas developing at the site of vascular prosthesis, and that the rate of these sarcoma, associated with foreign bodies, was much higher than the rate of sarcomas in general. All sarcomas associated with Dacron grafts were high grade histology and disseminated at the time of presentation. The authors also describe sarcoma reported at the site of other foreign bodies, such as shrapnel, bullets, steel plates and retained surgical sponges. They also note that the latency period from the acquisition of the foreign body and the development of sarcoma had a mean of 33 years. They document that a chronic foreign body reaction, the same "microscopic foreign body reaction" described by Dr. David Robinson in his Sept 2013 deposition as being clinically insignificant, was the etiology of this carcinogenesis. The authors also describe sarcomas developing in rodents after inert plastic polymers were placed in their soft tissue: "The sarcomas developed in rodents in which thick fibrous capsules developed around the implanted material." The authors conclude: "For unknown reasons, the cells in this inflammatory and repair process may undergo a malignant transformation, probably associated with oncogene activation and tumor suppressor gene inactivation. Further studies are warranted to search for the mechanisms involved in foreign

²⁵⁹ Kwon S., et al, Female Pelvic Med Reconstruct Surg, Volume 18, Number 4, July/August 2012

²⁶⁰ Ahuja, S., et al, Gynecol Surg (2011) 8:217-221

²⁶¹ Moller, K., et al, Gynecologic Oncology 94 (2004) 840-842.

²⁶² Ben-Izhak, O., et al, Am J Surg Pathology, Issue: Volume 23 (11), 1999, p. 1418

body tumorigenesis." To date no manufacturer of mesh products has investigated this oncogenic potential as the authors recommended.

In a report from the International Agency for Research on Cancer: Surgical Implants and Other Foreign Bodies, "When several polymers were tested in rats according to the same experimental protocol, sarcoma incidences ranged from 70% (polypropylene) to 7% (silicone)".²⁶³ "Polymeric implants prepared as thin smooth films (with the exception of poly(glycolic acid)) are POSSIBLY CARCINOGENIC TO HUMANS."²⁶⁴

Given the fact that hernia mesh placement increased in the 1990's with the advent of laparoscopic placement, and that vaginal mesh placed for SUI and POP accelerated in the 2000's, we may be on the cusp of an ever increasing number of foreign body tumors associated with vaginal mesh. To not inform doctors of this significant potential risk, and to not warn patients of an existent risk of cancer, when known, is unconscionable.

Ethicon did not undertake any long term testing to determine whether or not these warnings on the polypropylene resin manufacturers MSDS were associated with long term consequences for permanent human use. This is true despite the fact that Ethicon has knowledge of three of these cancer reports (Kwon, Moller and Ahuja) as they are referenced in Ethicon's 2013 Clinical Evaluation Report regarding TVT.²⁶⁵

Additionally, there is no evidence that Ethicon made any effort to inform surgeons of important information contained in various Manufacturer Safety Data Sheets (MSDS) regarding the use of polypropylene. This information includes the dangers of using polypropylene in a permanent implanted medical device. And, that laboratory studies on rats showed that polypropylene caused sarcomas in laboratory rats. Clearly, these facts are critical information

²⁶³ International Agency for Research on Cancer, Summaries and Evaluations, Vol.:74 (1999).

²⁶⁴ McGregor, D.B., et al, European Journal of Cancer 36 (2000) 307-313 (emphasis added).

²⁶⁵ Eth.Mesh.10150515

relevant to both the surgeon evaluating his or her treatment options and to the patient's informed consent decisions. As a result, Ethicon failed to act like a reasonable and prudent medical device manufacturer.

G. Ethicon did not properly inform physicians and their patients that toxicity testing of the polypropylene mesh revealed that it was cytotoxic or toxic to cells.

Cytotoxicity means toxicity to the cells causing cell injury or death.²⁶⁶ In a May 26, 2000 Ethicon Memo titled "Review of biocompatibility on the tension-free vaginal tape (TVT) system for compliance to FDA,"²⁶⁷ the review contains a "Cytotoxicity Risk Assessment for the TVT (Ulmsten) Device" from August 8, 1997.²⁶⁸ The Cytotoxicity Assessment states "there is some evidence to suggest that the PP [polypropylene] mesh from the sterile Ulmsten device may have cytotoxic potential."²⁶⁹ In addition, ISO Elution testing "resulted in marked cytotoxicity in tests conducted at Ethicon (Scotland)."²⁷⁰

According to former Ethicon Medical Director, Dr. David Robinson, Ethicon never performed "a single long-term study . . . to determine whether or not the Ethicon mesh is clinically cytotoxic in women."²⁷¹ In addition, in its IFUs and Patient Brochures, Ethicon never informed physicians or their patients about the possibility of cytotoxicity.²⁷² Dr. Robison testified that if there is a clinical related outcome related to cytotoxicity, it is reasonable for physicians to want to know that the mesh in the TVT product had been tested multiple times to be severely or marked cytotoxic.²⁷³

²⁶⁶ Robinson, 9/11/13 1091:11-21.

²⁶⁷ Eth.Mesh.06852118-06852129 (5/26/2000 Biocompatibility Review).

²⁶⁸ Eth.Mesh.06852120 (8/8/1997 Cytotoxicity Risk Assessment).

²⁶⁹ *Id.* and Robinson 9/11/13, 1099:4-9

²⁷⁰ Robinson,, 9/11/13,1095:2-21

²⁷¹ Robinson, 9/11/13 , 1101:24-1102:4

²⁷² Robinson, 9/11/13, 1114:7-18

²⁷³ Robinson, 9/11/13, 1115:13-19

Cytotoxicity can cause death to cells that can lead to an inflammatory response leading to a multitude of injuries, including serious adverse complications such as erosions, chronic pelvic pain, recurrence, worsening incontinence, dyspareunia, wound infection, rejection of the mesh, sexual dysfunction, urinary and defecatory dysfunction or the need for additional surgeries. Ethicon did not undertake any long term testing to determine whether the marked cytotoxicity found in the TVT mesh had long term consequences for permanent human use. This is true despite the fact that its own test results showed the mesh to be cytotoxic.

The potential for cytotoxicity or cell death is important information that physicians need to know in order to pass the information on to their patients so that an informed decision can be made about whether to have a permanent medical device implanted in their body. It is clear from Ethicon's Medical Director David Robinson that this information was never passed on to physicians despite the fact that it would have been reasonable for physicians to have this information. As a result, Ethicon did not act as a reasonably prudent medical device manufacturer in it failed to inform physicians and their patients about the risk of its mesh being cytotoxicity.

H. Ethicon's promotional materials sent to physicians related to TVT were inaccurate and failed to reveal material facts about complications and conflict of interests regarding data promoted in the materials.

Since the TVT was first launched, Ethicon has sent materials in various forms to physicians promoting long term follow up data on the original cohort of patients implanted with the TVT from 1995-1996²⁷⁴. In addition, the materials tout low complication rates related to various adverse reactions, including erosions. These materials include press releases, marketing brochures and email blasts.

²⁷⁴ Eth.Mesh.0015598; Eth.Mesh.00658058; Eth.Mesh.01186068; Eth.Mesh.02236784; Eth.Mesh.02237103; Eth.Mesh.03459211; Eth.Mesh. 05183409; Eth.Mesh.00339437; Eth.Mesh.05794787.

The long term data primarily relied on by Ethicon throughout these materials relates to the Ulmsten/Nilsson studies. These studies were originally started by Dr. Ulmsten, the inventor of the TVT, and continued by Dr. Nilsson after Dr. Ulmsten's death. Prior to selling the TVT to Johnson & Johnson, Dr. Ulmsten owned a company called Medscand. As discussed more fully below, Johnson & Johnson hired Dr. Ulmsten and Medscand to conduct studies related to the TVT. To this day, Ethicon relies heavily on these studies and uses them in numerous promotional materials despite the fact that Ethicon never disclosed to physicians the potential conflict of interest and inherent bias that exists due to Dr. Ulmsten's relationship with Ethicon and Johnson & Johnson. In addition, Ethicon never disclosed to physicians that the device used in the original Medscand study was different than the TVT device. It is important to physicians using the TVT that the data in these types of promotional materials is accurate, unbiased and that physicians are informed about any potential conflicts of interest in the data contained within the materials. In other words, physicians rely on Ethicon to provide fair and balanced information and to ensure that physician have been given all the data and not just the positive press release data.

Despite using the Ulmsten data to promote the TVT, Ethicon never disclosed to physicians the bias and inherent conflict of interest related to the Ulmsten data. Specifically, in its promotional materials, Ethicon (Johnson and Johnson) never informed physicians about its relationship and contracts with Professor Ulmsten and his company Medscand. It is clear from the contracts that the publications and data from Dr. Ulmsten were contracted for hire by Johnson and Johnson International.²⁷⁵

²⁷⁵ Eth.Mesh.08696085-134

The License and Supply Agreement between Johnson and Johnson International and Medscand (Ulmsten's Company) dated February 13, 1997, states In section 3.6 Milestone Payments Johnson and Johnson International (JJI) shall pay shall pay to Medscand the following payments (b). A payment in the amount of \$400,000.00 due on February 28, 1997; provided, however, that in the event that Clinical Trials as specified in Exhibit C have not been completed by such date, then such amount shall not be due until the completion of the Clinical Trials.²⁷⁶

Under Exhibit F Professor Ulmsten Consulting Agreement with Professor Alf Ivar Ulmsten, under section 4 Confidential Information: Rights to inventions and Copyrights (B) it states any copyrightable work whether published or unpublished created by supplier Dr. Ulmsten directly as a result of or during the performance of services herein shall be considered a work made for hire, to the fullest extent permitted by law and all rights, titles and interest herein, including worldwide copyrights shall be the property of the company as the employer and party specially commissioned said work.²⁷⁷

Finally, Exhibit C Clinical Trials, it states the results of clinical trials will be considered acceptable if, first, they do not differ significantly from the results published in the original article published in the Int. Urogynecol J 1996-7:81-86 by U.Ulmsten et al., with regards to the following items: Safety 1.1, per-operative complications 1.2 , post-operative complications 1[(year from operation)] 2. Efficacy. Second Long term results [over 1 year from operation] do not show a deterioration of rates significantly different from those of the standard suburethral slingplasties. It is assumed that from 12 – 60 months a gradual decrease in efficacy of 5% is normal. Third. No significant numbers of unexpected i.e. not addressed in the original article published in the Int. Urogynecol J 1996-781-86 by U.Ulmsten et al. procedure related i.e. not

²⁷⁶ Eth.Mesh.08696091

²⁷⁷ Eth.Mesh.08696116

addressed in the review article published in the Int. Urogynecol J 1994-5: 228-239 by G. N. Ghomiem et.al. complications appear at any time in the postoperative course.²⁷⁸

In total, Dr. Ulmsten stood to gain millions of dollars for the 6 papers that he published on the TVT device. In addition, the results of those studies would be found acceptable for payment because they did not differ from the parameters sent by Johnson & Johnson regarding complications and efficacy. The Ulmsten studies have an inherent conflict of interest and bias as they were “made for hire” and standards were set by Johnson & Johnson. As set forth above, if Dr. Ulmsten did not meet the standards set forth by Johnson & Johnson, he did not receive substantial payments for the “studies.” As a result of this relationship, there is a clear conflict of interest and potential for enormous bias issues.

The conflict of interest and bias created by the relationship between Ethicon and Dr. Ulmsten was acknowledged by Dr. Axel Arnaud, Ethicon’s European Medical Director, in a recent deposition. Specifically, Dr. Arnaud testified that such an agreement like the one discussed above between Dr. Ulmsten and Johnson & Johnson creates a potential conflict of interest.²⁷⁹ Dr. Arnaud also acknowledged that when Johnson & Johnson enters into this type of agreement with a physician or his company and the study is published, there “certainly” needs to be a disclosure of the relationship.²⁸⁰

Additionally, Former Ethicon Medical Director, Dr. David Robinson, testified that in his experience working in the industry for medical device manufacturers, it is best that potential biases be disclosed.²⁸¹ He also testified that if publications from somebody like Ulmsten or Nilsson about safety and efficacy are being published, it is best if they disclose that they have a

²⁷⁸ Eth.Mesh.08696132.

²⁷⁹ Arnaud 7/20/13 497:24-501:21, 509:8-17

²⁸⁰ Arnaud 7/20/13 514:17-515:1

²⁸¹ Robinson, 9/11/2013 1021:21-1022:1

financial bias or conflict of interest.²⁸² In fact, in an April 2009 email exchange with Medical Director Piet Hinoul about a physician who, like Ulmsten, is a consultant and inventor for competitor Boston Scientific, Dr. Robinson states that that situation presents “enormous bias issues.”²⁸³ Despite two of its medical directors testifying that the relationship between Ulmsten and carried over to Nilsson presents a conflict of interest and bias, Ethicon has never disclosed this information in its promotional pieces. This is information physicians and patients have a right to know so that a proper informed decision regarding the value of the data in the studies and the use of the product can be made.

Interestingly, Peter Cecchini, regulatory fellow at Ethicon and the person responsible for the 510K submission of the TVT did not know about the financial incentives or the terms of the agreements between Dr. Ulmsten and Ethicon.²⁸⁴ Mr. Cecchini testified that although the regulations at the time did not require Ethicon to disclose these biases, current regulations would require Ethicon to inform the FDA when submitting a 510K about these kinds of potential biases.²⁸⁵ In addition, Mr. Cecchini testified that despite the FDA regulations in place at the time of the 510K submission for TVT, Johnson & Johnson’s Credo requires that that type of information be “passed on.”²⁸⁶ Mr. Cecchini testified that if he would have known about the potential bias of the Ulmsten studies today, he would have passed it on.²⁸⁷

Aside from never disclosing to physicians the underlying conflict of interest and bias of the Ulmsten studies in its promotional pieces, Ethicon also never informed them about other problems with the data, including incomplete data on the original cohort, data incorrectly reported and erosion rates underreported. In the original 510k submission for TVT Classic,

²⁸² Robinson, 9/11/2012 1022:14-20

²⁸³ Eth.Mesh.03259439; Robinson 9/11/13 1021:21-1022:22

²⁸⁴ Cecchini, 10/22/13, 36:24-37:8

²⁸⁵ Cecchini, 10/22/13, 37:9-43:4

²⁸⁶ Cecchini, 10/22/13, 37:9-43:4

²⁸⁷ Id.

Ethicon used Medscand data from the Scandinavian Multicenter Study.²⁸⁸ The report shows that 12 month follow was obtained for 90 of the original 131 patients, without explanation of why there was a loss of 41 patients from the study. The study also describes complication of wound infection: "while the vaginal infection required surgical intervention with resection of exposed mesh". This represents a vaginal mesh erosion/extrusion/ exposure and needs to be reported as such. However, when the paper was published (Ulmsten, Int Urogynecol J 1998), the paper states that there was no defect healing and no tape rejections. It further misrepresents the outcome for this patient as "The patient with the wound infection had vaginal atrophy. After minimal vaginal wall resection and effective local estrogen treatment she healed without further intervention. There was no tape rejection."

If Ulmsten had reported a mesh erosion/extrusion/exposure with mesh excision in his study, it would not have been acceptable under Exhibit C of his consulting contract for payment of the \$400,000.²⁸⁹ This demonstrates that the results of this paper were potentially biased by the payment Ulmsten would receive for favorable data and should discount the data. At the very least, Ethicon should have informed physicians about the relationship between Ethicon and the Ulmsten studies.

Many of the marketing brochures tout the "[t]he urethral erosion rate less than or equal to that of traditional slings; no reported urethral erosions in 10 studies of 50+ patients."²⁹⁰ The reference used for the first part of this statement is from Dr. Gary Leach who looked at traditional sling procedures done before 1993, when traditional slings were performed at the bladder neck and purposely placed under tension to treat severe stress urinary incontinence from intrinsic sphincter deficiency (particularly among Urogynecologists).

²⁸⁸ Eth.Mesh 00371587

²⁸⁹ Eth.Mesh 08696132

²⁹⁰ Eth.Mesh 00339439

The second part of this statement regarding “no ureteral erosions” is incorrect. In published studies, Dr. Karram found one case of urethral erosion in his study of 350 Gynecare TVTs performed (Karram Obstet. Gynecol 2003) and Hammad found nine cases of urethral erosion in his study (Hammad Eur Urol 2005).²⁹¹ His study followed the complications of 1459 patients 993 of whom had Gynecare TVT, while the remainder has SPARC procedures. While the authors do not break down the incidence of urethral erosion by product, it is exceedingly unlikely that all erosions happen in the SPARC group.

The statement regarding “no ureteral erosions” also did not include deTayrac's 2003 paper of 61 patients (31 TVTs) which showed a 3% urethral erosion rate.²⁹² Dr. Shlomo Raz described a study of 26 patients who presented with voiding dysfunction, including symptoms of severe urethral, pelvic and genital pain, urinary retention, recurrent UTIs, de-novo urgency with urge incontinence found to have mesh from a sling procedure in the bladder or urethra.²⁹³ Their patients were found to have been treated conservatively with anticholinergic medication. They conclude that "dysfunctional voiding symptoms after sling procedure should elicit a high degree of suspicion if pharmacotherapy is not successful in alleviating symptoms...Cystoscopy should be considered if the patient remains symptomatic despite pharmacotherapy".

In one of the Nilsson studies, Dr. Nilsson describes four patients on "anticholinergics" (Int Urogynecol J 2008 Table 3). They conclude "It is also encouraging to see that no late adverse effects of the polypropylene tape material was found and that erosion of the tape into adjacent tissue did not occur." However, this statement cannot be made for 4 patients who are on pharmacotherapy without a cystoscopy, which was not performed in the 11 year follow-up study.

²⁹¹ Karram MM, Segal JL, Vassallo BJ and Kleeman SD, “Complications and untoward effects of the tension-free vaginal tape procedure,” *Ob & Gyn*, 2003 (101:929-32).

²⁹² de Tayrac R, et al, , “A prospective randomized trial comparing tension-free vaginal tape for surgical treatment of stress urinary incontinence,” *Am J Obstet Gynecol*, 2004 (190:602-8).

²⁹³ Deng DY, Rutman M, Raz S and Rodriguez LV, “Presentation and management of major complications of midurethral slings: Are complications under reported,” *Neurourology Urodynamics*, , 2007(26:46-52).

Dr. Raz's review of the literature found multiple cases of urethral erosions in a large series with TVT.²⁹⁴ There have also been multiple case reports attesting to the fact that urethral erosion does occur specifically with Gynecare TVT products.²⁹⁵ To imply that urethral erosion does not occur is not giving physicians fair and balance information about the true incidence of urethral erosions with TVT products.

Later, Nilsson publishes the 5 year follow-up of this cohort.²⁹⁶ He describes the cohort: "a prospective open multicenter trial was conducted in the Nordic countries at the beginning of 1995. The short-term results were published in 1998." This implies that these are the same patients as published in 1998. It is interesting or an incredible coincidence that the exact number of patients receiving 12 months of follow-up in the Medscand publication (90) was the exact number being described in the 5 year study. There is again no mention of the outcome of the other 41 patients from the original cohort. Another interesting detail in the 5 year study is that the original number of centers used for the study (6) was now down to 3, again without explanation. The 5 year report does describe the original patient with the wound infection but again fails to mention she had mesh excised, "1 case (1.1%) of infection of operating site was observed."

In 2006, Dr. Nilsson published a different study on long term outcome of patients with TVT²⁹⁷. He describes his new patient population: "A multi-center study comprising only carefully selected primary cases revealed a promising cure rate of 85% after 5 years (reference

²⁹⁴ *Karram 2003, Hammad 2005*

²⁹⁵ Sweat, S., et al, "Polypropylene Mesh Tape for Stress Urinary Incontinence: Complication of Urethral Erosion and Outlet Obstruction," *J Urology*, 2002 (168:144-146).; Gerstenbluth, RE. et al, "Simultaneous Urethral Erosion of Tension-Free Vaginal Tape and Woven Polyester Pubovaginal Sling," *J Urol*, 2003(2 Pt 1) (170:525-6); , Vassallo, BJ., et al, "Management of Iatrogenic Vaginal Constriction," *Am J Obstet Gynecol*, 2003 (192(3):512-20); Haferkamp, A., et al, "Urethral Erosion of Tension-Free Vaginal Tape," *J Urol* 2002 (167(1): 250).

²⁹⁶ Ulmsten data; Nilsson *Int Urogynecol J* 2001.

²⁹⁷ Kuuva, N, et al. "Long-term results of the tension-free vaginal tape operation in an unselected group of 129 stress incontinent women," *Acta Obstetricia Gynecologica Scandinavica* 2006(85: 4 482-87)

his 5 year study) and 81% at 7 years."²⁹⁸ These two papers are the subject of many press releases and marketing brochures, but they never described that these were carefully selected patients.

"To our knowledge, the long-term effect and effectiveness of the TVT procedure has not yet been studied in an unselected patient group. We earlier reported 16-month follow-up results of a general patient group referred to a tertiary medical unit and comprising primary, recurrent, mixed, and low pressure urethra cases. In the present study, we report the long-term results in the same above-mentioned group." They describe a 3.1% mesh "visualized" rate, half of which needed surgical resection. These results, more representative of what one would see in a normal practice, is never mentioned in press releases or marketing documents.

Conversely, when Ethicon receives adverse information, it does not make it into the promotional pieces. Dr. AC Wang's abstract, "Tension-Free Vaginal Tape (TVT) for Urinary Stress Incontinence - A Preliminary Report" was used in the original 510(k) submission in October of 1997 as support for FDA clearance of the TVT.²⁹⁹ However, when Dr. Wang reported that he had 25 cases of "failure of vaginal healing considered by him to be potential tape rejection...in each case the revision failed within 2 weeks, requiring further surgery to excise mesh and repair the vaginal wound," this important information never made it into the marketing materials or press releases.³⁰⁰

The long-term follow-up data (Ulmsten/Nilsson data) used by Ethicon to promote the lack of risk of TVT is spurious at best. We have incomplete data on the original cohort, data that is falsely reported, original sites that were excluded without explanation, and a lead investigator who had a significant relationship and financial incentive to reach certain results with the data. In fact, Ethicon's Medical Director Dr. Isenberg believes the Ulmsten data is "narrow and

²⁹⁸ Nilsson Obstet Gynecol 2004

²⁹⁹ Eth.Mesh.00371551

³⁰⁰ Eth.Mesh.00409675.

uncontrolled.”³⁰¹ This is the same data which is now used repeatedly in promotional and marketing materials sent to physicians.

I. Ethicon’s Patient Brochures misstate information regarding complications and success rates and lack fair balance.

In its patient brochures, Ethicon routinely represents to patients and physicians that the success rate of the TVT device is between 97 and 98%. These claims are based on the 5 year, 7 year, and 11 year follow-up to the Ulmsten/Nilsson study. What is not adequately explained to patients is that this “success rate” is a combination of patients who are “cured” and those who are considered “improved,” a fact that is not disclosed in advertisements directed at patients, but is generally disclosed in advertisements directed at doctors. Further, Ethicon does not disclose to patients or physicians that the numerous authors of this study on which Ethicon has made the cornerstone of its marketing program are paid consultants of Ethicon, including, Dr. Ulmsten, Dr. Nilsson, Dr. Falconer, and Dr. Rezapour.³⁰² Ethicon also does not inform patients of the existence of other long-term studies which show a much lower success rate for TVT.

Ethicon’s Patient brochures tell patients that 98% of women treated with the Gynecare TVT are still dry or report significantly less leakage after 7 years.³⁰³ This claim is combined with an assertion that the TVT is trusted by over 1 million women. This claim undoubtedly gives women the false impression that this claim of 98% “success” with TVT is based on all of the more than 1 million women who have been treated with TVT, when in fact this statement is based on a single study of 90 patients, only 80 of which were available for follow-up after 7

³⁰¹ Eth.Mesh 08793648

³⁰² Eth.Mesh.09746615; Eth.Mesh.09748842; Eth.Mesh.09748848; Eth.Mesh.08696050; Eth.Mesh.08167644.

³⁰³ Eth.Mesh .00163583

years.³⁰⁴ The brochure also does not tell patients that the actual cure rate in this study was only 81.3%. with the TVT until page 11 of the brochure, and then only in fine print at the bottom of the brochure. The brochure also does not inform patients of what the criteria for a patient to be considered “significantly improved” is. Further, patients were not informed that 8% of these 80 patients felt their incontinence had become worse between their 5 year evaluation, and the 7 year evaluation that was the basis of this study. Ethicon also knew of other studies showing a much lower success rate than the single 90 patient study they had elected to feature in their patient brochures. For example one study found only a 63% success rate after only two years, and that study was sponsored by Ethicon and was a randomized controlled trial (RCT), a study type known to produce the highest quality evidence.³⁰⁵ Yet, Ethicon chose not to share that information with patients. By not presenting patients with an overview of the likely success rates with TVT and instead of selecting a single study which happened to have a high apparent success rate when the “substantially improved” patients were included as a success, Ethicon failed to include fairly balanced material in their brochure.

This same patient brochure also informs women who are considering having TVT surgery that “few patients experience complications.”³⁰⁶ This is a clear misrepresentation of data and complication rates known to Ethicon and the scientific community. For example, in one two-year study, 39% of the patients who received a TVT device experienced a least one complication. In fact, Ethicon’s medical director, Dr. Weisberg testified that he knew bladder perforations with TVT occurred in 2 to 3 percent of patients ranging up to 19 percent in some

³⁰⁴ Nilsson, CG. et. al., “Seven year follow-up of the Tension –Free Vaginal Tape procedure for treatment of urinary incontinence,” *Ob & Gyn*, 2004 (104: S5-S8)

³⁰⁵ Ward K et. al., “Prospective multicentre randomized trial of tension-free vaginal tape and colposuspension as primary treatment for stress incontinence,” *BMJ* 2002 (325:1-7).

³⁰⁶ Eth.Mesh 00163583

populations.³⁰⁷ In 2002, Ethicon was also aware of a report of 25 wound healing defects out of approximately 600 patients – all were suspected to be tape rejections by a very experienced TVT surgeon. This would represent an erosion rate of slightly over 4% seen in these 600 patients.³⁰⁸ The *Barber* study found an erosion rate in patients receiving the TVT device of between 5 and 6%,³⁰⁹ a rate consistent with what the President of Ethicon, Renee Selman, believed to be the overall rate of erosions seen with the TVT (“between five and ten percent”). The complication rates known to Ethicon are not consistent with the claim in its patient brochures that few women experience complications, and grossly understate the risk to women deciding whether or not to have the procedure.³¹⁰

Some patient brochures also contain statements and taglines that further give the patient the false impression that complications are rare, and minimize the invasiveness, recovery time and potential complications with the procedure. For example, tag lines like “[s]top coping, start living,”³¹¹ and “[o]ne day you have stress urinary incontinence, the next day you don’t- end of story,”³¹² mislead patients into thinking that they will never again have to deal with the symptoms of stress urinary incontinence or other urinary symptoms when the clinical studies show that anywhere between 19% and 37% of patients in clinical studies still have some stress urinary symptoms after treatment with TVT. Further, it minimizes the patients’ consideration of the possibility that they may have new symptoms to manage as a result of their operation, including but not limited to potential recurrent urinary tract infections, new urge incontinence, dyspareunia, chronic pain, erosions, and urinary retention.

³⁰⁷ Weisberg 5/ 31,13422:20-423:1.

³⁰⁸ Id. at 434:1-437:25; Eth.Mesh.00409674.

³⁰⁹ Barber et al., “Transobturator tape compared with tension-free vaginal tape for the treatment of stress urinary incontinence,” *Obstet Gynecol*, 2008(111:611-621).

³¹⁰ Selman 6/21/13, 583:17-584:17.

³¹¹ Eth.Mesh.08003279.

³¹² Eth.Mesh.08003263.

According to Associate Medical Director Dr. Meng Chen, the TVT family of product Patient Brochures also call the procedure “minimally invasive.”³¹³ This is true despite the fact that patients have reported that the procedure was “not minimally invasive.”³¹⁴ Dr. Chen agreed that she had received reports that patients’ lives were ruined despite that they were told the procedure was minimally invasive.³¹⁵ Dr. Chen also testified that she categorized the TVT procedure as “majorly invasive” when comparing it with surgical sutures.³¹⁶ This is conjunction with Ethicon’s definition of “Major Invasive Surgeries” as outlined in its internal documents.³¹⁷ In short, Ethicon internally calls the TVT procedure “majorly invasive” but markets it to patients as minimally invasive.

Other statements in the brochure give women a false impression of the recovery time. “Recovery is quick,”³¹⁸ and “Short recovery period and quick return to normal activities,”³¹⁹ give patients a false impression of the actual recovery time involved. It is not unusual for women to take 4 weeks or longer before they can safely return to work after the TVT procedure, particularly in jobs that require physical exertion as part of the job functions. Further, the recovery time varies from patient to patient, with some patients taking 8 weeks or even longer to completely heal and return to normal activities.

J. Ethicon’s Collection and Reporting of Adverse Events and Complications to Physicians and Patients Is Misleading, Inaccurate and Incomplete.

Ethicon did not actively try to determine how many patients were hurt by its devices, including the TVT, or how severely they were hurt. Instead, Ethicon had a “passive” system of

³¹³ Chen, 10/30/13, 262:10-21

³¹⁴ Chen, 10/30/13, 266:19-267:3

³¹⁵ Chen, 10/30/13, 267:5-13

³¹⁶ Chen, 10/30/13, 265:23-266:11

³¹⁷ Eth.Mesh.00321804

³¹⁸ Eth.Mesh.08003291.

³¹⁹ Eth.Mesh.08003264.

measuring how many and what type of adverse events the TVT was causing. Ethicon's Director of Post-Marketing Surveillance testified that this type of passive collecting of reports understates how many people are actually being hurt by its devices:

THE WITNESS: So we -- from a reactive perspective for complaints, we can only process the complaints that are reported to us, so -- and as we discussed earlier, they come from many different avenues; but again, they're reactive in nature, which means we are processing what is given to us or reported to us.

....

Q. You understand that spontaneous adverse event reporting, such as your department collects and analyzes, has been demonstrated to substantially underquantify the real complications in the world?

A: So the adverse events that are reported to us, complications, complaints that are reported to us, are a subset of the events, complaints, complications that occur in the field.³²⁰

In fact, Ethicon employees ensured that they would not "actively" collect any complaints. When discussing how to perform a marketing survey with a number of physicians Dan Smith, a TVT Project Leader, wanted to ensure Ethicon people did not ask physicians questions that might "collect" a complaint:

Just a thought with regard to us collecting information. Paul, what was the ruling from our compliance group regarding us asking questions/collecting data, did we have to log issues as complaints???? et cetera. If so, we should do this in a manner that avoids this issue.³²¹

Dr. David Robinson, Ethicon's Medical Director, noted a reason that Ethicon might not want to actively collect adverse events about its products: "[I]f this starts getting reported, it is going to scare the daylights out of docs."³²²

³²⁰ Lamont Dep. (4/4/13) 389:25-390:23; Yale Dep. (8/7/13) 126:20-127:7 ("So you would agree that generally in a passive complaint collection, which is what Ethicon had prior to this discussion about the registry, for example, in a passive collection, that it is well known and well recognized that adverse events are underreported. Correct? THE WITNESS: In general, the basic understanding in the world of complaints and adverse events is that you do not get 100 percent reporting, that, you know, it is not the perfect collection model to gather. So, yes, they are, in some manner, underreported.").

³²¹ ETH.MESH.01811770.

³²² ETH.MESH.00756984 (Email from David Robinson, M.D. to Giselle Bonet and Marty Weisberg).

Even though Ethicon limited its “surveillance” to passively collecting complaints, it did not do this well. For example, Mark Yale, the head of Ethicon’s Worldwide Customer Quality team testified that all Ethicon employees had a legal duty to report any and all complaints to the Company about which they became aware.³²³ When shown documentation, Yale admitted that this collection system was flawed. For example, employees in a US call center failed to report complaints,³²⁴ employees in Eastern Europe did not know they were required to inform the Company of complaints and adverse events,³²⁵ one Portuguese employee testified that he would not have reported the complaint, but someone had already informed the regulatory authorities:

Q. So Francisco in Portugal working for Johnson & Johnson Medical says he wouldn't have reported this to you, this complication, except for the fact that somebody reported it to their regulatory authorities. Right?

A. That's what he wrote. Correct.³²⁶

This line of questioning led to a consistent theme about adverse events and complications tracking at Ethicon – you don’t know what you don’t know. Yale testified:

Q. So as you sit here today, you have no idea how many other complaints didn't make it here from Portugal, because Francisco Noronha from Johnson & Johnson decided that if it wasn't reported to his regulatory agency, he's not going to tell you about it. Right?

THE WITNESS: I don't know what I don't know.³²⁷

When David Menneret, an employee of the mesh manufacturer at Ethicon SARL received a complaint about mesh being frayed (a significant issue as discussed above) he was unsure whether to report it as a “complaint” into the Ethicon complaint tracking system. He wrote:

Please see attached below a letter...regarding Mesh fraying. I don't know exactly who should be informed of this kind of customer feeling so feel free to forward to anyone concerned. Do you think this should be entered as a complaint in the system?³²⁸

³²³ Yale Dep. (8-7-2013) 140:12 to 140:16.

³²⁴ Yale Dep. (8-7-2013) 145:12 to 145:15.

³²⁵ Yale Dep. (8-7-2013) 155:21 to 155:25.

³²⁶ Yale Dep. (8-7-2013) 159:5 to 159:10.

³²⁷ Yale Dep. (8-7-2013) 160:16 to 160:24.

³²⁸ ETH.MESH.01814252.

Again, Yale testified that he could not know how many complaints went to the manufacturer about the fraying from the manufacturing process that ultimately were not reported to Ethicon's complaint tracking system. He testified as follows:

Q. You don't know how many times Menneret didn't report a complaint either. Right? You don't know what you don't know. Right?

THE WITNESS: As I said before, I do not know what I do not know....³²⁹

Prior to March of 2006, Ethicon did not even have a formal procedure in place to capture adverse events from its own clinical trials. Therefore, it had no idea how many adverse events occurred but were not reported from those trials.³³⁰

In addition to the marketing materials, Ethicon also provided physicians with "Complications Statements" during training or upon request. These "Complication Statements" relied upon the information captured in Ethicon's complaint system – the same system described above. Accordingly, the capture of information for these statements was already severely compromised. However, even for those events Ethicon did capture, the reporting of these events in the Complications Statements was completely misleading.

Joseph Scavona, a complaint analyst, was responsible for creating one of these Complications Statements that was provided to physicians. He described how he created the statement and how, if a woman had multiple injuries, he only listed one injury on the chart. He wrote:

[S]ome complaints could be described with multiple main & sub categories, but each complaint was only labeled with one of these categories (e.g. patient had pain, bleeding, hematoma, exposure, and dyspareunia thus complaint was coded only "mesh exposure").³³¹

³²⁹ Yale Dep. (8-7-2013) 168:24 to 169:12.

³³⁰ Yale Dep. (8-7-2013) 194:22 to 195:7.

³³¹ ETH.MESH.02122904 (Ex. 970) (Email from Joseph Scavona to others re "TVT Complications Statement 2008"). Complications Statement attached at ETH.MESH.00007091 at 2 (Ex. T-970).

This completely misrepresented the actual harms data. Moreover, the person making these decision, Scavona, was not a medical doctor. He recognized these limitations and requested that medical review the complications data, but it did not occur.³³² Instead, physicians were provided with misleading, inaccurate and incomplete information in the Complications Statements.³³³

In my opinion Ethicon's collection and reporting of adverse events and complications to physicians and patients was incomplete, inaccurate and misleading. As manufacturers are the only entities with access to complaint information, physicians and patients must rely upon them to provide timely, accurate and complete information. Ethicon failed to do so. Without accurate information, physicians could not and cannot obtain informed consent from their patients, nor can patients give informed consent. Ethicon's complaint collecting and reporting system made this impossible.

K. The Benefits of the TVT Are Outweighed By the Severe, Debilitating and Life Changing Complications Associated With the TVT

It is my opinion, based on my training, experience and extensive review of the literature and Ethicon's internal documents that the benefits of the TVT are outweighed by the severe, debilitating and life changing complications associated with the medical device. It is clear that a substantial number of women who are implanted with the TVT have already and will continue to suffer chronic, debilitating erosions or pain, among other complications, and these life changing complications outweigh the benefits of the TVT, a device used to treat a quality of life issue. This is especially true given that traditional surgeries like the Burch and pubovaginal slings are not associated with the frequency or extent of these life changing complications.

³³² Id.

³³³ Yale Dep. (8-8-2013) 294 to 300.

Ethicon senior medical directors admit that the efficacy of the TVT is equivalent to the traditional surgeries like the Burch.³³⁴ I agree. Traditional surgeries are not associated with TVT mesh based complications like contraction and erosion, however, with clinically significant erosion. And, further, although traditional surgeries can cause symptoms such as pain following surgery, including dyspareunia, the risk, duration, extent and severity of chronic pain including dyspareunia following the TVT is much greater than with traditional surgeries, and of course those surgeries do not result in the often untreatable complications and symptoms that result from the TVT mesh.

Unfortunately, although there have been a large number of studies and publications involving the TVT over the years, the quality of most of the studies is not good, and the amount of bias included in the studies and publications adds to the limited value that the studies offer about long term, severe and debilitating complications like chronic pain and erosions associated with the TVT. The most recent Cochrane review of mid-urethral slings, Ogah (2011), concluded that most trials involving mid-urethral slings had short follow-up and the quality of evidence was variable such that the quality of evidence for the majority of trials was moderate with a minority having low-to-moderate evidence.³³⁵ Few trials reported outcomes after 1 year and long term adverse effects had yet to be determined. There are only a handful of RCTs involving the TVT that are long term, and major and long term complications would unlikely be picked up in these RCTs in part because they are designed with a primary endpoint of efficacy, not safety. The true incidence are more likely to be determined by registries or databases, but published registries do not track certain complications such as pain or dyspareunia, and have not been designed to monitor long term problems (Tamussino, 2001 and 2007; Kuuva 2002, Collinet

³³⁴ Robinson 9/11/13 939:21-940:5

³³⁵ Ohah, et. al., Minimally Invasive Synthetic Suburethral Sling Operations for Stress Urinary Incontinence in Women: A Short Version Cochrane Review. *Neurology and Urodynamics* 30:284-291 (2011).

2008, Dykorn 2010). This void in studying and presenting the true incidence and nature of long term and life altering complications, along with the biases inherent in many of the studies, and other factors, negates the value of the large majority of the studies, and as a result, other sources of data such as published case series are relevant and important to truly understand the nature of these complications. Ethicon's internal documents and data, which are not publically available, present a very different picture of the TVT than the information that has been shared with patients and physicians.

I have done an in-depth review and analysis of the studies, and am prepared to discuss the studies including the small number of studies that have tracked chronic pain, dyspareunia and erosions on a long term basis. The Abbott study is particularly noteworthy, however. Abbott (2014) described a series of 347 patients evaluated for mesh related complications from 2006 - 2010. Approximately 50% had a sling only and an additional 26% had a sling and TVM mesh. The median time from placement to evaluation was 5.8 months with a range of 0 – 65.2. This would mean that many of these patients would not have been captured in registries or RCT's with one year or less follow-up. Also only 26% were seen by another facility before attending one of the study sites, meaning that at least $\frac{3}{4}$ of these complications were not known to the implanting physician, again highlighting the limited utility of data at the primary site. The authors found 30% of patients had dyspareunia, 43% had erosion and 35% had pelvic pain.³³⁶ This study highlights the degree and severity of the complications that mesh slings like the TVT are causing and, importantly, that physicians in the real world simply do not have the information about the severity of the problem. This is why it is extremely important for

³³⁶ Abbott, et. al., Evaluation and management of complications from synthetic mesh after pelvic reconstructive surgery: a multicenter study. American Journal of Obstetrics & Gynecology. (Feb. 2014).

manufacturers of slings like Ethicon to accurately and fully report the risks and complications associated with the mesh devices to doctors – something Ethicon simply has not done.

V. CONCLUSION

Ethicon has marketed and sold the TVT despite the fact that it contains numerous characteristics that make it unsuitable for implantation in a woman's vagina. These characteristics include the following: (1) degradation of the mesh; (2) small pores and heavy weight; (3) chronic foreign body reaction; (4) fraying and particle loss; (5) Infections and Bio-films; (6) roping and curling of the mesh; (7) loss of pore size with tension; (8) fibrotic bridging leading to scar plate formation and mesh encapsulation; and (9) shrinkage/contraction of the encapsulated mesh.

Not only does Ethicon sell a product which should never be put in the vagina, it failed to inform physicians and their patients about numerous risks associated with the product despite the fact that they knew of all of these risks before the product was launched. This failure by Ethicon has robbed women of their ability to make a proper informed decision with their physician about whether or not to have a permanent medical device implanted into their body. In addition, despite having knowledge to the contrary, Ethicon never told physicians and their patients that the TVT was associated with cancer and could be toxic to their bodies. Finally, while keeping this information from women, Ethicon marketed its product with promotional pieces that did not disclose key conflict of interest information or the true complication rates of its products.


As a result of these failures as fully set forth in this report, the TVT has caused and will continue to cause a multitude of injuries in women, including the possibility of multiple erosions that can occur throughout one's lifetime, chronic and debilitating pelvic pain, recurrence, worsening incontinence, dyspareunia that can be chronic, wound infection, rejection of the mesh,

sexual dysfunction, urinary and defecatory dysfunction, vaginal scarring, wound healing problems, injury to ureters, pelvic abscess formation, risk of infection, and/or the need for additional surgeries, among others.

All opinions I have are to a reasonable degree of medical certainty. I understand discovery is still ongoing in this case and I reserve my right to amend my opinions if further information is provided in any form including, but not limited to corporate documents, depositions and the expert reports of both Plaintiff and Defense experts.

I declare under penalty of perjury that the foregoing is true and correct.

This 9th day of June 2014



Bruce Rosenzweig, M.D.